

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4861

Client: AECOM

Analytical Method: 8270E DataFile: BF140596.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB165052BS	Pyridine	50	49.3	ug/L	99		*		29	97	
	1,4-Dichlorobenzene	50	46.6	ug/L	93				76	103	
	2-Methylphenol	50	49.4	ug/L	99				69	109	
	3+4-Methylphenols	50	49.2	ug/L	98				67	106	
	Hexachloroethane	50	46.3	ug/L	93				76	118	
	Nitrobenzene	50	44.6	ug/L	89				58	106	
	Hexachlorobutadiene	50	44.9	ug/L	90				69	101	
	2,4,6-Trichlorophenol	50	48.0	ug/L	96				61	110	
	2,4,5-Trichlorophenol	50	47.2	ug/L	94				70	106	
	2,4-Dinitrotoluene	50	48.7	ug/L	97				60	115	
	Hexachlorobenzene	50	46.6	ug/L	93				73	106	
	Pentachlorophenol	100	100	ug/L	100				47	114	