

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4861 SAS No.: P4861 SDG NO.: P4861

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/18/2024

Lab File ID: BF140456.D Time Analyzed: 16:14

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	499315	11.398	237967	14.045	262467	15.533
UPPER LIMIT	998630	11.898	475934	14.545	524934	16.033
LOWER LIMIT	249658	10.898	118984	13.545	131234	15.033
EPA SAMPLE NO.						
01 WC-11-A-202411	538236	11.40	315994	14.05	154098	15.53

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.