

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4861

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VN084919.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1118WBS01	Vinyl chloride	20	20.6	ug/L	103			65	117	
	1,1-Dichloroethene	20	20.9	ug/L	104			74	110	
	2-Butanone	100	120	ug/L	120			65	122	
	Carbon Tetrachloride	20	21.6	ug/L	108			77	113	
	Chloroform	20	22.0	ug/L	110			79	113	
	Benzene	20	21.2	ug/L	106			82	109	
	1,2-Dichloroethane	20	21.9	ug/L	110			80	115	
	Trichloroethene	20	20.3	ug/L	102			77	113	
	Tetrachloroethene	20	20.4	ug/L	102			67	123	
	Chlorobenzene	20	21.5	ug/L	108			82	109	