

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4861

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VN084923.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1118WBSD01	Vinyl chloride	20	18.2	ug/L	91	12		65	117	20
	1,1-Dichloroethene	20	18.1	ug/L	91	13		74	110	20
	2-Butanone	100	100	ug/L	100	18		65	122	20
	Carbon Tetrachloride	20	18.5	ug/L	93	15		77	113	20
	Chloroform	20	19.0	ug/L	95	15		79	113	20
	Benzene	20	17.9	ug/L	90	16		82	109	20
	1,2-Dichloroethane	20	18.6	ug/L	93	17		80	115	20
	Trichloroethene	20	17.5	ug/L	88	15		77	113	20
	Tetrachloroethene	20	17.1	ug/L	86	17		67	123	20
	Chlorobenzene	20	17.7	ug/L	89	19		82	109	20