

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: P4861 SAS No.: P4861 SDG NO.: P4861
 Lab File ID: VN084911.D Date Analyzed: 11/18/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	168638	8.22	277153	9.09	245519	11.87
UPPER LIMIT	337276	8.718	554306	9.594	491038	12.365
LOWER LIMIT	84319	7.718	138577	8.594	122760	11.365
EPA SAMPLE NO.						
WC-11-A-202411	175303	8.22	312575	9.10	285945	11.87
VN1118WBL01	175680	8.22	312901	9.10	269131	11.87
VN1118WBS01	184217	8.22	316400	9.09	286519	11.87
VN1118WBSD01	187181	8.22	326112	9.10	290218	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.