

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: P4866 SAS No.: P4866 SDG No.: P4866

Instrument ID: BNA_N Calibration Date/Time: 11/21/2024 23:01

Lab File ID: BN035235.D Init. Calib. Date(s): 11/13/2024 11/13/2024

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:40 16:15

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.713	0.581		-18.5	50.0
Fluoranthene-d10	1.225	0.987		-19.4	50.0
2-Fluorophenol	1.015	1.087		7.1	50.0
Phenol-d6	1.273	1.273		0.0	50.0
Nitrobenzene-d5	0.348	0.189		-45.7	50.0
2-Fluorobiphenyl	1.624	0.255		-84.3	50.0
2,4,6-Tribromophenol	0.289	0.242		-16.3	50.0
Terphenyl-d14	0.839	6.857		717.3	50.0
1,4-Dioxane	0.363	0.381		5.0	50.0

All other compounds must meet a minimum RRF of 0.010.