

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112124\
 Data File : BN035235.D
 Acq On : 21 Nov 2024 23:01
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Nov 21 23:33:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 14:00:03 2024
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.322	152	3350	0.400	ng	0.00
7) Naphthalene-d8	10.073	136	7674	0.400	ng	0.00
13) Acenaphthene-d10	13.976	164	4706	0.400	ng	-0.01
19) Phenanthrene-d10	16.746	188	9785	0.400	ng	# 0.00
29) Chrysene-d12	21.008	240	924	0.400	ng	# 0.00
35) Perylene-d12	23.092	264	8785	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	4.975	112	3641	0.428	ng	0.00
5) Phenol-d6	6.528	99	4263	0.400	ng	0.00
8) Nitrobenzene-d5	8.461	82	1602	0.240	ng	0.00
11) 2-Methylnaphthalene-d10	11.674	152	4455	0.326	ng	0.00
14) 2,4,6-Tribromophenol	15.494	330	1140	0.336	ng	0.00
15) 2-Fluorobiphenyl	12.608	172	1202	0.063	ng	0.00
27) Fluoranthene-d10	18.802	212	9657	0.322	ng	0.00
31) Terphenyl-d14	19.429	244	6336	3.268	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.011	88	1275	0.419	ng	92
3) n-Nitrosodimethylamine	3.307	42	994	0.351	ng	# 88
6) bis(2-Chloroethyl)ether	6.773	93	3222	0.403	ng	97
9) Naphthalene	10.116	128	7849	0.392	ng	100
10) Hexachlorobutadiene	10.415	225	1778	0.303	ng	# 99
12) 2-Methylnaphthalene	11.750	142	5199	0.352	ng	96
16) Acenaphthylene	13.688	152	7880	0.392	ng	99
17) Acenaphthene	14.040	154	5046	0.383	ng	98
18) Fluorene	15.045	166	6711	0.346	ng	99
20) 4,6-Dinitro-2-methylph...	15.142	198	322	0.158	ng	# 47
21) 4-Bromophenyl-phenylether	15.952	248	2324	0.373	ng	# 84
22) Hexachlorobenzene	16.051	284	2690	0.416	ng	98
23) Atrazine	16.250	200	1260	0.225	ng	95
24) Pentachlorophenol	16.411	266	1189	0.393	ng	97
25) Phenanthrene	16.784	178	10390	0.403	ng	99
26) Anthracene	16.883	178	9521	0.403	ng	99
28) Fluoranthene	18.834	202	12485	0.353	ng	# 97
30) Pyrene	19.201	202	12814	4.164	ng	99
32) Benzo(a)anthracene	21.026	228	11745	3.650	ng	96
33) Chrysene	21.026	228	11745	3.685	ng	99
34) Bis(2-ethylhexyl)phtha...	21.026	149	1040	0.616	ng	# 85
36) Indeno(1,2,3-cd)pyrene	25.145	276	13630	0.389	ng	# 90
37) Benzo(b)fluoranthene	22.475	252	11553	0.391	ng	# 95
38) Benzo(k)fluoranthene	22.519	252	11920	0.403	ng	# 94
39) Benzo(a)pyrene	23.002	252	10296	0.396	ng	# 93
40) Dibenzo(a,h)anthracene	25.162	278	10355	0.373	ng	# 90
41) Benzo(g,h,i)perylene	25.762	276	11294	0.382	ng	# 91


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112124\
 Data File : BN035235.D
 Acq On : 21 Nov 2024 23:01
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Quant Time: Nov 21 23:33:46 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 14:00:03 2024
 Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.322 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC





Tgt Ion:152 Resp: 3350
 Ion Ratio Lower Upper
 152 100
 150 155.4 124.5 186.7
 115 62.1 47.8 71.6




#2
 1,4-Dioxane
 Concen: 0.419 ng
 RT: 3.011 min Scan# 18
 Delta R.T. -0.000 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01



Tgt Ion: 88 Resp: 1275
 Ion Ratio Lower Upper
 88 100
 43 22.0 16.9 25.3
 58 56.2 39.0 58.4









#3
n-Nitrosodimethylamine
Concen: 0.351 ng
RT: 3.307 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 994
Ion Ratio Lower Upper
42 100
74 166.1 120.8 181.2
44 10.4 2.9 4.3#



#4
2-Fluorophenol
Concen: 0.428 ng
RT: 4.975 min Scan# 290
Delta R.T. -0.007 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion: 112 Resp: 3641
Ion Ratio Lower Upper
112 100
64 48.9 39.7 59.5
63 26.9 23.0 34.4









#5
Phenol-d6
Concen: 0.400 ng
RT: 6.528 min Scan# 505
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 4263
Ion Ratio Lower Upper
99 100
42 14.5 11.4 17.0
71 35.9 34.6 51.8



#6
bis(2-Chloroethyl)ether
Concen: 0.403 ng
RT: 6.773 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion: 93 Resp: 3222
Ion Ratio Lower Upper
93 100
63 61.9 47.4 71.2
95 31.4 26.2 39.4





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.073 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:136 Resp: 7674





Ion Ratio Lower Upper

136 100

137 12.2 9.8 14.8

54 5.6 4.0 6.0

68 6.0 4.2 6.2



#8

Nitrobenzene-d5

Concen: 0.240 ng

RT: 8.461 min Scan# 766

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion: 82 Resp: 1602


Ion Ratio Lower Upper

82 100

128 49.0 36.5 54.7

54 37.5 28.7 43.1





#9

Naphthalene

Concen: 0.392 ng

RT: 10.116 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:128 Resp: 7849

Ion Ratio Lower Upper

128 100

129 11.9 9.6 14.4

127 14.5 11.6 17.4



#10

Hexachlorobutadiene

Concen: 0.303 ng

RT: 10.415 min Scan# 949

Delta R.T. -0.011 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:225 Resp: 1778


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 65.5 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.326 ng
RT: 11.674 min Scan# 1116
Delta R.T. -0.004 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 4455
Ion Ratio Lower Upper
152 100
151 20.6 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 11.750 min Scan# 1116
Delta R.T. -0.004 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:142 Resp: 5199
Ion Ratio Lower Upper
142 100
141 90.4 70.1 105.1
115 40.1 34.4 51.6





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 13.976 min Scan# 14

Delta R.T. -0.011 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:164 Resp: 4706

Ion Ratio Lower Upper

164 100

162 107.5 82.2 123.2

160 54.6 40.2 60.4



#14

2,4,6-Tribromophenol

Concen: 0.336 ng

RT: 15.494 min Scan# 1613

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:330 Resp: 1140


Ion Ratio Lower Upper

330 100

332 0.0 76.9 115.3#

141 34.2 29.6 44.4









#15
2-Fluorobiphenyl
Concen: 0.063 ng
RT: 12.608 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 1202
Ion Ratio Lower Upper
172 100
171 37.4 28.2 42.4
170 24.8 19.0 28.6



#16
Acenaphthylene
Concen: 0.392 ng
RT: 13.688 min Scan# 1444
Delta R.T. -0.011 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:152 Resp: 7880
Ion Ratio Lower Upper
152 100
151 20.4 15.8 23.8
153 13.3 10.2 15.4









#17
Acenaphthene
Concen: 0.383 ng
RT: 14.040 min Scan# 1496
Delta R.T. -0.011 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 5046
Ion Ratio Lower Upper
154 100
153 115.6 91.6 137.4
152 59.8 45.8 68.6



#18
Fluorene
Concen: 0.346 ng
RT: 15.045 min Scan# 1571
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:166 Resp: 6711
Ion Ratio Lower Upper
166 100
165 100.4 79.1 118.7
167 13.8 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.746 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 9785
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.1 4.3 6.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.158 ng
RT: 15.142 min Scan# 1580
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:198 Resp: 322
Ion Ratio Lower Upper
198 100
51 63.3 29.9 44.9#
105 64.2 23.7 35.5#









#21
4-Bromophenyl-phenylether
Concen: 0.373 ng
RT: 15.952 min Scan# 1650
Delta R.T. -0.000 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 2324
Ion Ratio Lower Upper
248 100
250 97.8 81.2 121.8
141 65.5 29.1 43.7#



#22
Hexachlorobenzene
Concen: 0.416 ng
RT: 16.051 min Scan# 1658
Delta R.T. -0.012 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:284 Resp: 2690
Ion Ratio Lower Upper
284 100
142 35.7 28.2 42.4
249 30.9 26.2 39.2





#23

Atrazine

Concen: 0.225 ng

RT: 16.250 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:200 Resp: 1260

Ion Ratio Lower Upper

200 100

173 33.4 22.6 33.8

215 48.2 37.8 56.6



#24

Pentachlorophenol

Concen: 0.393 ng

RT: 16.411 min Scan# 1687

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:266 Resp: 1189


Ion Ratio Lower Upper

266 100

264 63.5 49.4 74.0

268 63.4 52.6 79.0





#25

Phenanthrene

Concen: 0.403 ng

RT: 16.784 min Scan# 1725

Delta R.T. -0.012 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:178 Resp: 10390

Ion Ratio Lower Upper

178 100

176 19.8 15.2 22.8

179 15.6 12.6 18.8



#26

Anthracene

Concen: 0.403 ng

RT: 16.883 min Scan# 1725

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:178 Resp: 9521


Ion Ratio Lower Upper

178 100

176 18.7 14.6 22.0

179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.322 ng
RT: 18.802 min Scan# 1919
Delta R.T. -0.005 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 9657
Ion Ratio Lower Upper
212 100
106 12.4 6.0 9.0#
104 7.3 3.5 5.3#



#28
Fluoranthene
Concen: 0.353 ng
RT: 18.834 min Scan# 1926
Delta R.T. -0.005 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:202 Resp: 12485
Ion Ratio Lower Upper
202 100
101 9.9 5.2 7.8#
203 17.0 13.8 20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.008 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 924

Ion Ratio Lower Upper

240 100

120 29.2 3.8 5.6#

236 40.2 22.2 33.2#



#30

Pyrene

Concen: 4.164 ng

RT: 19.201 min Scan# 2005

Delta R.T. -0.005 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:202 Resp: 12814


Ion Ratio Lower Upper

202 100

200 21.3 16.7 25.1

203 17.7 14.2 21.4









#31
 Terphenyl-d14
 Concen: 3.268 ng
 RT: 19.429 min Scan# 2087
 Delta R.T. -0.009 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 6336
 Ion Ratio Lower Upper
 244 100
 212 10.6 7.6 11.4
 122 14.4 6.1 9.1#



#32
 Benzo(a)anthracene
 Concen: 3.650 ng
 RT: 21.026 min Scan# 2396
 Delta R.T. -0.009 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01

Tgt Ion:228 Resp: 11745
 Ion Ratio Lower Upper
 228 100
 226 30.5 21.4 32.0
 229 19.8 15.7 23.5









#33
 Chrysene
 Concen: 3.685 ng
 RT: 21.026 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Tgt Ion:228 Resp: 11745




Ion	Ratio	Lower	Upper
228	100		
226	30.5	23.7	35.5
229	19.8	16.2	24.4




#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.616 ng
 RT: 21.026 min Scan# 2396
 Delta R.T. -0.000 min
 Lab File: BN035235.D
 Acq: 21 Nov 2024 23:01

Tgt Ion:149 Resp: 1040

Ion	Ratio	Lower	Upper
149	100		
167	20.9	23.2	34.8#
279	0.0	3.2	4.8#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.092 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 8785
Ion Ratio Lower Upper
264 100
260 27.6 20.9 31.3
265 60.1 35.4 53.2#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.389 ng
RT: 25.145 min Scan# 3487
Delta R.T. -0.012 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Tgt Ion:276 Resp: 13630
Ion Ratio Lower Upper
276 100
138 26.1 13.0 19.4#
277 23.9 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.391 ng

RT: 22.475 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN035235.D


Acq: 21 Nov 2024 23:01

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:252 Resp: 11553

Ion Ratio Lower Upper

252 100

253 24.8 19.3 28.9

125 14.0 6.2 9.2#



#38

Benzo(k)fluoranthene

Concen: 0.403 ng

RT: 22.519 min Scan# 2589

Delta R.T. -0.009 min

Lab File: BN035235.D

Acq: 21 Nov 2024 23:01




Tgt Ion:252 Resp: 11920


Ion Ratio Lower Upper

252 100

253 25.0 19.5 29.3


125 15.4 6.5 9.7#







#39
Benzo(a)pyrene
Concen: 0.396 ng
RT: 23.002 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 10296
Ion Ratio Lower Upper
252 100
253 26.4 20.1 30.1
125 16.3 7.5 11.3#



#40
Dibenzo(a,h)anthracene
Concen: 0.373 ng
RT: 25.162 min Scan# 3493
Delta R.T. -0.015 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01


Tgt Ion:278 Resp: 10355
Ion Ratio Lower Upper
278 100
139 20.7 9.3 13.9#
279 26.2 19.7 29.5





#41
Benzo(g,h,i)perylene
Concen: 0.382 ng
RT: 25.762 min Scan# 30
Delta R.T. -0.015 min
Lab File: BN035235.D
Acq: 21 Nov 2024 23:01

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:	276	Resp:	11294
Ion	Ratio	Lower	Upper
276	100		
277	24.5	19.9	29.9
138	24.5	11.6	17.4#

