

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: P4866

SDG NO.: P4866

Lab File ID: BN035218.D

DFTPP Injection Date: 11/21/2024

Instrument ID: BNA_N

DFTPP Injection Time: 12:43

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	23.4
68	Less than 2.0% of mass 69	0.0 (0.0) 1
69	Mass 69 relative abundance	30.3
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	40
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.5
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	11.3
442	Greater than 50% of mass 198	68.4
443	15.0 - 24.0% of mass 442	12.7 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN035219.D	11/21/2024	13:22
PB165012BL	PB165012BL	BN035220.D	11/21/2024	13:58
RW7-SP200-20241114	P4866-01	BN035223.D	11/21/2024	15:48
RW7-SP201-20241114	P4866-02	BN035225.D	11/21/2024	17:00
PB165012BS	PB165012BS	BN035226.D	11/21/2024	17:36
PB165012BSD	PB165012BSD	BN035227.D	11/21/2024	18:12
SSTDCCC0.4EC	SSTDCCC0.4	BN035235.D	11/21/2024	23:01