

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P4868

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓

The Initial Calibration met the requirements .

The Continuous Calibration File ID BN035219.D met the requirements except for 2-
Fluorobiphenyl, Fluoranthene-d10, Nitrobenzene-d5 and Terphenyl-d14 which are not
our target compound, therefore no corrective action taken.

The Continuous Calibration File ID BN035235.D met the requirements except for 2-
Fluorobiphenyl and Terphenyl-d14 which are not our target compound, therefore no
corrective action taken.

6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓
7.	Surrogate Recoveries Meet Criteria		✓

If not met, list those compounds and their recoveries which fall outside the acceptable
ranges.

The Surrogate recoveries met the acceptable criteria except for
RW5-SP100-20241114 [Nitrobenzene-d5 - 46%, Terphenyl-d14 - 441%],
RW5-SP201-20241114 [Terphenyl-d14 - 3736%],
RW5-SP303-20241114 [2-Fluorobiphenyl - 38%, Terphenyl-d14 - 677%],
PB165012BL [2-Fluorobiphenyl - 42%, Terphenyl-d14 - 323%],
PB165012BS [2-Fluorobiphenyl - 22%, Terphenyl-d14 - 556%],
PB165012BSD [2-Fluorobiphenyl - 19% and Terphenyl-d14 - 567%], failing surrogates
was not associated with client parameter list, therefore no corrective action taken.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

	NA	NO	YES
8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.			
The Blank Spike met requirements for all samples . The Blank Spike Duplicate met requirements for all samples .			
9. Internal Standard Area/Retention Time Shift Meet Criteria		✓	
Comments: The Internal Standards Areas met the acceptable requirements except for RW5-SP303-20241114, RW5-SP201-20241114, PB165012BS and PB165012BSD, failing Internal Standards were not associated with client parameter list, therefore no corrective action taken.			
10. Extraction Holding Time Met			✓
If not met, list number of days exceeded for each sample:			
11. Analysis Holding Time Met			✓
If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

Samples RW5-SP100-20241114 was diluted at 5X. This sample initially analyzed straight in sequence BN112024 but END CCAL is missing therefore based on that result lab analyzed this sample directly with 5X while Straight analysis given as screening data in Miscellaneous section.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

QA REVIEW

Date