

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4868
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4868 **SAS No.:** P4868
Matrix: Water **Sample ID:** P4867-04 **Client ID:** TT-072-IDWGW-20241113DUP
Percent Solids for Sample: NA **Duplicate ID** P4867-04DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	0.20	U	0.20	U			CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4868
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4868 **SAS No.:** P4868
Matrix: Water **Sample ID:** P4867-04MS **Client ID:** TT-072-IDWGW-20241113MSD
Percent Solids for Sample: NA **Duplicate ID** P4867-04MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Mercury	ug/L	20	4.34		3.99		8		CV

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4868
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4868 **SAS No.:** P4868
Matrix: Water **Sample ID:** P4868-03 **Client ID:** RW5-SP303-20241114DUP
Percent Solids for Sample: NA **Duplicate ID** P4868-03DUP **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	50.0	U	50.0	U			P
Antimony	ug/L	20	25.0	U	25.0	U			P
Arsenic	ug/L	20	10.0	U	10.0	U			P
Barium	ug/L	20	50.0	U	50.0	U			P
Beryllium	ug/L	20	3.00	U	3.00	U			P
Cadmium	ug/L	20	3.00	U	3.00	U			P
Calcium	ug/L	20	4720		4740		0		P
Chromium	ug/L	20	4.33	J	4.42	J	2		P
Cobalt	ug/L	20	1.62	J	1.70	J	5		P
Copper	ug/L	20	10.0	U	10.0	U			P
Iron	ug/L	20	46.5	J	44.5	J	4		P
Lead	ug/L	20	6.00	U	6.00	U			P
Magnesium	ug/L	20	1780		1790		1		P
Manganese	ug/L	20	36.2		36.6		1		P
Nickel	ug/L	20	2.38	J	2.38	J	0		P
Potassium	ug/L	20	792	J	807	J	2		P
Selenium	ug/L	20	10.0	U	10.0	U			P
Silver	ug/L	20	5.00	U	5.00	U			P
Sodium	ug/L	20	10700		10900		2		P
Thallium	ug/L	20	20.0	U	20.0	U			P
Vanadium	ug/L	20	20.0	U	20.0	U			P
Zinc	ug/L	20	9.05	J	9.29	J	3		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: Tetra Tech NUS, Inc. **Level:** LOW **SDG No.:** P4868
Contract: TETR06 **Lab Code:** CHEM **Case No.:** P4868 **SAS No.:** P4868
Matrix: Water **Sample ID:** P4868-03MS **Client ID:** RW5-SP303-20241114MSD
Percent Solids for Sample: NA **Duplicate ID** P4868-03MSD **Percent Solids for Spike Sample:** NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Aluminum	ug/L	20	1070		1070		0		P
Antimony	ug/L	20	416		414		0		P
Arsenic	ug/L	20	403		402		0		P
Barium	ug/L	20	103		102		1		P
Beryllium	ug/L	20	99.6		98.7		1		P
Cadmium	ug/L	20	98.0		97.0		1		P
Calcium	ug/L	20	5170		5110		1		P
Chromium	ug/L	20	218		217		0		P
Cobalt	ug/L	20	101		101		0		P
Copper	ug/L	20	158		157		1		P
Iron	ug/L	20	1560		1570		1		P
Lead	ug/L	20	475		470		1		P
Magnesium	ug/L	20	2770		2750		1		P
Manganese	ug/L	20	137		136		1		P
Nickel	ug/L	20	249		247		1		P
Potassium	ug/L	20	5780		5810		1		P
Selenium	ug/L	20	1010		1010		0		P
Silver	ug/L	20	31.9		32.0		0		P
Sodium	ug/L	20	11800		12000		2		P
Thallium	ug/L	20	1030		1050		2		P
Vanadium	ug/L	20	159		157		1		P
Zinc	ug/L	20	115		116		1		P