

Initial and Continuing Calibration Blank Summary

Client: AECOM

SDG No.: P4871

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133475

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	0.001	0.0025	J	0.00099	0.005	11/15/2024
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/15/2024
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/15/2024
Sample ID: CCB3 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/15/2024
Sample ID: CCB4 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00099	0.005	11/15/2024

Initial and Continuing Calibration Blank Summary

Client: AECOM

SDG No.: P4871

Project: Meeker Ave Plumes Superfund Site RI FS

RunNo.: LB133498

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/18/2024
Sample ID: CCB1 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/18/2024
Sample ID: CCB2 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/18/2024
Sample ID: CCB3 Cyanide	mg/L	< 0.0025	0.0025	U	0.00099	0.005	11/18/2024