

Laboratory Control Sample Summary

Client: AECOM

SDG No.: P4871

Project: Meeker Ave Plumes Superfund Site RI FS

Run No.: LB133498

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB165017BS							
Cyanide	mg/L	0.1	0.095		95	1	85-115	11/18/2024