

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** P4871
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4871 **SAS No.:** P4871

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165049BS							
Arsenic	ug/L	4000	4010		100	80 - 120	P
Barium	ug/L	1000	1070		107	80 - 120	P
Cadmium	ug/L	1000	1020		102	80 - 120	P
Chromium	ug/L	2000	2120		106	80 - 120	P
Lead	ug/L	5000	5120		102	80 - 120	P
Selenium	ug/L	10000	10100		101	80 - 120	P
Silver	ug/L	380	390		103	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: AECOM **SDG No.:** P4871
Contract: AECO02 **Lab Code:** CHEM **Case No.:** P4871 **SAS No.:** P4871

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB165080BS Mercury	ug/L	4.0	3.49		87	80 - 120	CV