

### CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

Continuing Calib Date: 11/15/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 20:40 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1242-1 (1)   | 5.92       | 5.92      | 5.82      | 6.02  | 0.00       |
| Aroclor-1242-2 (2)   | 5.95       | 5.95      | 5.85      | 6.05  | 0.00       |
| Aroclor-1242-3 (3)   | 6.01       | 6.01      | 5.91      | 6.11  | 0.00       |
| Aroclor-1242-4 (4)   | 6.11       | 6.11      | 6.01      | 6.21  | 0.00       |
| Aroclor-1242-5 (5)   | 6.84       | 6.84      | 6.74      | 6.94  | 0.00       |
| Tetrachloro-m-xylene | 4.76       | 4.76      | 4.66      | 4.86  | 0.00       |
| Decachlorobiphenyl   | 10.68      | 10.67     | 10.57     | 10.77 | -0.01      |

### CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

Continuing Calib Date: 11/15/2024 Initial Calibration Date(s): 11/08/2024 11/08/2024

Continuing Calib Time: 20:40 Initial Calibration Time(s): 02:45 10:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1242-1 (1)   | 5.16       | 5.16      | 5.06      | 5.26 | 0.00       |
| Aroclor-1242-2 (2)   | 5.18       | 5.18      | 5.08      | 5.28 | 0.00       |
| Aroclor-1242-3 (3)   | 5.36       | 5.36      | 5.26      | 5.46 | 0.00       |
| Aroclor-1242-4 (4)   | 5.44       | 5.44      | 5.34      | 5.54 | 0.00       |
| Aroclor-1242-5 (5)   | 5.97       | 5.97      | 5.87      | 6.07 | 0.00       |
| Tetrachloro-m-xylene | 4.05       | 4.05      | 3.95      | 4.15 | 0.00       |
| Decachlorobiphenyl   | 9.21       | 9.21      | 9.11      | 9.31 | 0.00       |