

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP068510.D Time Analyzed: 20:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.922	5.821	6.021	472.500	500.000	-5.5
Aroclor-1016-2	5.944	5.844	6.044	472.860	500.000	-5.4
Aroclor-1016-3	6.007	5.908	6.108	479.160	500.000	-4.2
Aroclor-1016-4	6.106	6.005	6.205	489.700	500.000	-2.1
Aroclor-1016-5	6.400	6.299	6.499	458.100	500.000	-8.4
Aroclor-1260-1	7.523	7.423	7.623	430.710	500.000	-13.9
Aroclor-1260-2	7.777	7.676	7.876	424.670	500.000	-15.1
Aroclor-1260-3	8.138	8.038	8.238	418.710	500.000	-16.3
Aroclor-1260-4	8.386	8.276	8.476	511.700	500.000	2.3
Aroclor-1260-5	8.714	8.613	8.813	416.630	500.000	-16.7
Decachlorobiphenyl	10.672	10.572	10.772	43.290	50.000	-13.4
Tetrachloro-m-xylene	4.758	4.657	4.857	49.020	50.000	-2.0

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 11/08/2024 11/08/2024

Client Sample No.: CCAL01 Date Analyzed: 11/15/2024

Lab Sample No.: AR1660CCC500 Data File : PP068510.D Time Analyzed: 20:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.156	5.058	5.258	493.060	500.000	-1.4
Aroclor-1016-2	5.176	5.079	5.279	483.510	500.000	-3.3
Aroclor-1016-3	5.356	5.259	5.459	468.000	500.000	-6.4
Aroclor-1016-4	5.396	5.298	5.498	477.230	500.000	-4.6
Aroclor-1016-5	5.615	5.517	5.717	486.010	500.000	-2.8
Aroclor-1260-1	6.658	6.560	6.760	461.540	500.000	-7.7
Aroclor-1260-2	6.844	6.747	6.947	449.250	500.000	-10.2
Aroclor-1260-3	7.002	6.905	7.105	455.620	500.000	-8.9
Aroclor-1260-4	7.476	7.378	7.578	461.480	500.000	-7.7
Aroclor-1260-5	7.715	7.617	7.817	440.080	500.000	-12.0
Decachlorobiphenyl	9.207	9.110	9.310	43.820	50.000	-12.4
Tetrachloro-m-xylene	4.050	3.951	4.151	50.120	50.000	0.2