

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110724

Review By	yogesh	Review On	11/11/2024 9:23:03 AM
Supervise By	Ankita	Supervise On	11/11/2024 9:48:52 AM
SubDirectory	PP110724	HP Acquire Method	HP Processing Method PP110724

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP068359.D	08 Nov 2024 02:13		YP\AJ	Ok
2	I.BLK	I.BLK	PP068360.D	08 Nov 2024 02:29		YP\AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP068361.D	08 Nov 2024 02:45		YP\AJ	Ok,M
4	AR1660ICC750	AR1660ICC750	PP068362.D	08 Nov 2024 03:02		YP\AJ	Ok
5	AR1660ICC500	AR1660ICC500	PP068363.D	08 Nov 2024 03:18		YP\AJ	Ok
6	AR1660ICC250	AR1660ICC250	PP068364.D	08 Nov 2024 03:34		YP\AJ	Ok,M
7	AR1660ICC050	AR1660ICC050	PP068365.D	08 Nov 2024 03:51		YP\AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP068366.D	08 Nov 2024 04:07		YP\AJ	Ok
9	AR1232ICC500	AR1232ICC500	PP068367.D	08 Nov 2024 04:23		YP\AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP068368.D	08 Nov 2024 04:40		YP\AJ	Ok
11	AR1242ICC750	AR1242ICC750	PP068369.D	08 Nov 2024 04:56		YP\AJ	Ok,M
12	AR1242ICC500	AR1242ICC500	PP068370.D	08 Nov 2024 05:12		YP\AJ	Ok
13	AR1242ICC250	AR1242ICC250	PP068371.D	08 Nov 2024 05:29		YP\AJ	Ok,M
14	AR1242ICC050	AR1242ICC050	PP068372.D	08 Nov 2024 05:45		YP\AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP068373.D	08 Nov 2024 06:01		YP\AJ	Ok
16	AR1248ICC750	AR1248ICC750	PP068374.D	08 Nov 2024 06:18		YP\AJ	Ok
17	AR1248ICC500	AR1248ICC500	PP068375.D	08 Nov 2024 06:34		YP\AJ	Ok
18	AR1248ICC250	AR1248ICC250	PP068376.D	08 Nov 2024 06:50		YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP110724

Review By	yogesh	Review On	11/11/2024 9:23:03 AM
Supervise By	Ankita	Supervise On	11/11/2024 9:48:52 AM
SubDirectory	PP110724	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP068377.D	08 Nov 2024 07:07		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP068378.D	08 Nov 2024 07:23		YPIAJ	Ok,M
21	AR1254ICC750	AR1254ICC750	PP068379.D	08 Nov 2024 07:39		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP068380.D	08 Nov 2024 07:56		YPIAJ	Ok,M
23	AR1254ICC250	AR1254ICC250	PP068381.D	08 Nov 2024 08:12		YPIAJ	Ok,M
24	AR1254ICC050	AR1254ICC050	PP068382.D	08 Nov 2024 08:28		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP068383.D	08 Nov 2024 08:45		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP068384.D	08 Nov 2024 09:01		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP068385.D	08 Nov 2024 09:17		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP068386.D	08 Nov 2024 09:34		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP068387.D	08 Nov 2024 09:50		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP068388.D	08 Nov 2024 10:39		YPIAJ	Ok,M
31	PP110724ICV500	ICVPP110724	PP068389.D	08 Nov 2024 10:55		YPIAJ	Ok,M
32	AR1242ICV500	ICVPP110724	PP068390.D	08 Nov 2024 11:12		YPIAJ	Ok,M
33	AR1248ICV500	ICVPP110724	PP068391.D	08 Nov 2024 11:28		YPIAJ	Ok,M
34	AR1254ICV500	ICVPP110724	PP068392.D	08 Nov 2024 11:45		YPIAJ	Ok,M
35	AR1268ICV500	ICVPP110724	PP068393.D	08 Nov 2024 12:01		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111524

Review By	yogesh	Review On	11/18/2024 10:57:56 AM
Supervise By	Ankita	Supervise On	11/18/2024 1:19:14 PM
SubDirectory	PP111524	HP Acquire Method	HP Processing Method PP110724

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,PP23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP068479.D	15 Nov 2024 09:06		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP068480.D	15 Nov 2024 09:22		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP068481.D	15 Nov 2024 09:38		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP068482.D	15 Nov 2024 09:55		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP068483.D	15 Nov 2024 10:11		YPIAJ	Ok,M
6	I.BLK	I.BLK	PP068484.D	15 Nov 2024 10:27		YPIAJ	Ok
7	PB164988BL	PB164988BL	PP068485.D	15 Nov 2024 11:28		YPIAJ	Ok
8	PB164988BS	PB164988BS	PP068486.D	15 Nov 2024 11:44		YPIAJ	Ok,M
9	P4858-01	BC258780-1-1	PP068487.D	15 Nov 2024 12:00		YPIAJ	Ok,M
10	P4858-02	BC258780-1-2	PP068488.D	15 Nov 2024 12:17		YPIAJ	Ok,M
11	P4858-03	BC258780-2-1	PP068489.D	15 Nov 2024 12:33		YPIAJ	Ok
12	P4858-04	BC258780-2-2	PP068490.D	15 Nov 2024 12:49		YPIAJ	Ok,M
13	P4858-05	HIH269B-1-1	PP068491.D	15 Nov 2024 13:06	AR1254 Hit	YPIAJ	Ok,M
14	P4858-06	HIH269B-1-2	PP068492.D	15 Nov 2024 13:22	AR1254 Hit	YPIAJ	Ok,M
15	P4858-09	BC225734-2-1	PP068493.D	15 Nov 2024 13:38		YPIAJ	Ok,M
16	P4858-11	BC248150-1-1	PP068494.D	15 Nov 2024 13:55		YPIAJ	Ok
17	AR1660CCC500	AR1660CCC500	PP068495.D	15 Nov 2024 15:14		YPIAJ	Ok
18	AR1242CCC500	AR1242CCC500	PP068496.D	15 Nov 2024 15:31		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111524

Review By	yogesh	Review On	11/18/2024 10:57:56 AM
Supervise By	Ankita	Supervise On	11/18/2024 1:19:14 PM
SubDirectory	PP111524	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PP068497.D	15 Nov 2024 15:47		YPIAJ	Ok
20	AR1254CCC500	AR1254CCC500	PP068498.D	15 Nov 2024 16:03		YPIAJ	Ok,M
21	I.BLK	I.BLK	PP068499.D	15 Nov 2024 16:20		YPIAJ	Ok
22	P4858-10	BC225734-2-2	PP068500.D	15 Nov 2024 16:36	DCB high in 2nd column	YPIAJ	Ok,M
23	P4858-07	BC225734-1-1	PP068501.D	15 Nov 2024 16:52		YPIAJ	Ok,M
24	P4858-08	BC225734-1-2	PP068502.D	15 Nov 2024 17:09		YPIAJ	Ok,M
25	P4858-12	BC248150-1-2	PP068503.D	15 Nov 2024 17:25		YPIAJ	Ok
26	P4858-13	Y2309-0044-1-1	PP068504.D	15 Nov 2024 17:42		YPIAJ	Ok,M
27	P4858-14	Y2309-0044-1-2	PP068505.D	15 Nov 2024 17:58		YPIAJ	Ok,M
28	P4858-15	BC274768-1-1	PP068506.D	15 Nov 2024 18:14		YPIAJ	Ok,M
29	P4858-16	BC274768-1-2	PP068507.D	15 Nov 2024 18:31		YPIAJ	Ok
30	P4858-17	BC274601-1-1	PP068508.D	15 Nov 2024 18:47		YPIAJ	Ok
31	P4858-18	BC274601-1-2	PP068509.D	15 Nov 2024 19:03		YPIAJ	Ok,M
32	AR1660CCC500	AR1660CCC500	PP068510.D	15 Nov 2024 20:24		YPIAJ	Ok
33	AR1242CCC500	AR1242CCC500	PP068511.D	15 Nov 2024 20:40		YPIAJ	Ok
34	AR1248CCC500	AR1248CCC500	PP068512.D	15 Nov 2024 20:57		YPIAJ	Ok
35	AR1254CCC500	AR1254CCC500	PP068513.D	15 Nov 2024 21:13		YPIAJ	Ok,M
36	I.BLK	I.BLK	PP068514.D	15 Nov 2024 21:29		YPIAJ	Ok
37	PB164990BL	PB164990BL	PP068515.D	15 Nov 2024 21:46		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP111524

Review By	yogesh	Review On	11/18/2024 10:57:56 AM
Supervise By	Ankita	Supervise On	11/18/2024 1:19:14 PM
SubDirectory	PP111524	HP Acquire Method	HP Processing Method PP110724
STD. NAME	STD REF.#		
Tune/Reschk	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775 PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773 PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB164990BS	PB164990BS	PP068516.D	15 Nov 2024 22:02		YPIAJ	Ok,M
39	PB164990BSD	PB164990BSD	PP068517.D	15 Nov 2024 22:18		YPIAJ	Ok,M
40	P4843-01	SW-WTS-01	PP068518.D	15 Nov 2024 22:35		YPIAJ	Ok,M
41	P4861-01	WC-11-A-202411	PP068519.D	15 Nov 2024 22:51		YPIAJ	Ok,M
42	P4867-02	TT-070-IDWGW-20241	PP068520.D	15 Nov 2024 23:08	DCB low in both column	YPIAJ	ReRun
43	P4867-03	TT-071-IDWGW-20241	PP068521.D	15 Nov 2024 23:24	DCB low in both column	YPIAJ	ReRun
44	P4867-04	TT-072-IDWGW-20241	PP068522.D	15 Nov 2024 23:40		YPIAJ	Ok
45	P4871-01	WC-10-A-202411	PP068523.D	15 Nov 2024 23:57		YPIAJ	Ok
46	AR1660CCC500	AR1660CCC500	PP068524.D	16 Nov 2024 00:45		YPIAJ	Ok
47	AR1242CCC500	AR1242CCC500	PP068525.D	16 Nov 2024 01:01		YPIAJ	Ok
48	AR1248CCC500	AR1248CCC500	PP068526.D	16 Nov 2024 01:18		YPIAJ	Ok
49	AR1254CCC500	AR1254CCC500	PP068527.D	16 Nov 2024 01:34		YPIAJ	Ok,M
50	I.BLK	I.BLK	PP068528.D	16 Nov 2024 01:51		YPIAJ	Ok

M : Manual Integration