

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093112.D Time Analyzed: 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.057	8.954	9.154	46.500	50.000	-7.0
Endrin	6.575	6.474	6.674	47.010	50.000	-6.0
gamma-BHC (Lindane)	4.327	4.228	4.428	50.430	50.000	0.9
Heptachlor	4.916	4.817	5.017	48.620	50.000	-2.8
Heptachlor epoxide	5.684	5.584	5.784	48.080	50.000	-3.8
Methoxychlor	7.501	7.399	7.599	46.550	50.000	-6.9
Tetrachloro-m-xylene	3.539	3.440	3.640	50.600	50.000	1.2

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL01 Date Analyzed: 11/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093112.D Time Analyzed: 13:32

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.916	7.816	8.016	49.310	50.000	-1.4
Endrin	5.640	5.541	5.741	54.960	50.000	9.9
gamma-BHC (Lindane)	3.610	3.511	3.711	54.390	50.000	8.8
Heptachlor	3.949	3.849	4.049	53.560	50.000	7.1
Heptachlor epoxide	4.731	4.632	4.832	54.300	50.000	8.6
Methoxychlor	6.614	6.515	6.715	51.800	50.000	3.6
Tetrachloro-m-xylene	2.777	2.678	2.878	52.890	50.000	5.8