

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093134.D Time Analyzed: 18:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.056	8.954	9.154	46.980	50.000	-6.0
Endrin	6.575	6.474	6.674	45.630	50.000	-8.7
gamma-BHC (Lindane)	4.328	4.228	4.428	50.460	50.000	0.9
Heptachlor	4.916	4.817	5.017	48.170	50.000	-3.7
Heptachlor epoxide	5.684	5.584	5.784	48.220	50.000	-3.6
Methoxychlor	7.500	7.399	7.599	42.500	50.000	-15.0
Tetrachloro-m-xylene	3.540	3.440	3.640	49.920	50.000	-0.2

CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No.: CCAL02 Date Analyzed: 11/18/2024

Lab Sample No.: PSTDCCC050 Data File : PL093134.D Time Analyzed: 18:21

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	7.915	7.816	8.016	51.160	50.000	2.3
Endrin	5.641	5.541	5.741	53.880	50.000	7.8
gamma-BHC (Lindane)	3.610	3.511	3.711	55.180	50.000	10.4
Heptachlor	3.949	3.849	4.049	53.770	50.000	7.5
Heptachlor epoxide	4.731	4.632	4.832	55.490	50.000	11.0
Methoxychlor	6.614	6.515	6.715	48.760	50.000	-2.5
Tetrachloro-m-xylene	2.778	2.678	2.878	53.760	50.000	7.5