

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: AECO02

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No. (PEM): PEM - PL093111.D Date Analyzed: 11/18/2024

Lab Sample No.(PEM): PEM Time Analyzed: 13:19

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.056	8.960	9.160	19.590	20.000	-2.1
Tetrachloro-m-xylene	3.539	3.490	3.590	21.050	20.000	5.3
alpha-BHC	3.994	3.940	4.040	10.790	10.000	7.9
beta-BHC	4.525	4.470	4.580	11.350	10.000	13.5
gamma-BHC (Lindane)	4.327	4.280	4.380	10.510	10.000	5.1
Endrin	6.573	6.500	6.640	42.900	50.000	-14.2
4,4'-DDT	7.024	6.950	7.090	86.130	100.000	-13.9
Methoxychlor	7.501	7.430	7.570	203.180	250.000	-18.7

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 10/28/2024 10/28/2024

Client Sample No. (PEM): PEM - PL093111.D Date Analyzed: 11/18/2024

Lab Sample No.(PEM): PEM Time Analyzed: 13:19

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.915	7.810	8.020	19.480	20.000	-2.6
Tetrachloro-m-xylene	2.777	2.730	2.830	20.570	20.000	2.9
alpha-BHC	3.279	3.230	3.330	9.860	10.000	-1.4
beta-BHC	3.909	3.860	3.960	10.850	10.000	8.5
gamma-BHC (Lindane)	3.609	3.560	3.660	9.470	10.000	-5.3
Endrin	5.640	5.570	5.710	49.940	50.000	-0.1
4,4'-DDT	6.038	5.970	6.110	106.700	100.000	6.7
Methoxychlor	6.614	6.540	6.680	238.650	250.000	-4.5