

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112524\
 Data File : BF140590.D
 Acq On : 25 Nov 2024 09:33
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 25 18:48:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
2	1,4-Dioxane	0.493	0.469	4.9	94	0.00
3	Pyridine	1.084	1.124	-3.7	92	0.00
4	n-Nitrosodimethylamine	0.637	0.569	10.7	86	-0.01
5 S	2-Fluorophenol	1.172	1.119	4.5	93	0.00
6	Aniline	1.091	1.234	-13.1	96	0.00
7 S	Phenol-d6	1.550	1.507	2.8	92	0.00
8	2-Chlorophenol	1.261	1.243	1.4	95	0.00
9	Benzaldehyde	0.820	0.779	5.0	103	0.00
10 C	Phenol	1.583	1.586	-0.2	93	0.00
11	bis(2-Chloroethyl)ether	1.211	1.180	2.6	93	0.00
12	1,3-Dichlorobenzene	1.417	1.382	2.5	97	0.00
13 C	1,4-Dichlorobenzene	1.435	1.399	2.5	96	0.00
14	1,2-Dichlorobenzene	1.344	1.313	2.3	95	0.00
15	Benzyl Alcohol	1.149	1.127	1.9	90	0.00
16	2,2'-oxybis(1-Chloropropane	1.431	1.423	0.6	95	0.00
17	2-Methylphenol	1.009	0.988	2.1	93	0.00
18	Hexachloroethane	0.536	0.520	3.0	95	0.00
19 P	n-Nitroso-di-n-propylamine	0.916	0.880	3.9	91	0.00
20	3+4-Methylphenols	1.298	1.252	3.5	91	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
22	Acetophenone	0.488	0.461	5.5	90	0.00
23 S	Nitrobenzene-d5	0.391	0.377	3.6	91	0.00
24	Nitrobenzene	0.404	0.386	4.5	91	0.00
25	Isophorone	0.652	0.631	3.2	90	0.00
26 C	2-Nitrophenol	0.179	0.180	-0.6	94	0.00
27	2,4-Dimethylphenol	0.214	0.218	-1.9	92	0.00
28	bis(2-Chloroethoxy)methane	0.397	0.394	0.8	94	0.00
29 C	2,4-Dichlorophenol	0.284	0.282	0.7	94	0.00
30	1,2,4-Trichlorobenzene	0.324	0.320	1.2	95	0.00
31	Naphthalene	1.030	1.017	1.3	95	0.00
32	Benzoic acid	0.164	0.179	-9.1	95	-0.01
33	4-Chloroaniline	0.308	0.331	-7.5	96	0.00
34 C	Hexachlorobutadiene	0.215	0.212	1.4	94	0.00
35	Caprolactam	0.088	0.090	-2.3	94	0.00
36 C	4-Chloro-3-methylphenol	0.318	0.314	1.3	91	0.00
37	2-Methylnaphthalene	0.654	0.650	0.6	94	0.00
38	1-Methylnaphthalene	0.641	0.637	0.6	94	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	91	0.00
40	1,2,4,5-Tetrachlorobenzene	0.586	0.586	0.0	94	0.00
41 P	Hexachlorocyclopentadiene	0.107	0.100	6.5	79	0.00
42 S	2,4,6-Tribromophenol	0.214	0.210	1.9	88	0.00
43 C	2,4,6-Trichlorophenol	0.367	0.368	-0.3	91	0.00
44	2,4,5-Trichlorophenol	0.399	0.405	-1.5	91	0.00
45 S	2-Fluorobiphenyl	1.342	1.315	2.0	92	0.00
46	1,1'-Biphenyl	1.493	1.483	0.7	93	0.00
47	2-Chloronaphthalene	1.131	1.133	-0.2	93	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112524\
 Data File : BF140590.D
 Acq On : 25 Nov 2024 09:33
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 25 18:48:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.363	0.361	0.6	89	0.00
49	Acenaphthylene	1.710	1.718	-0.5	94	0.00
50	Dimethylphthalate	1.317	1.305	0.9	91	0.00
51	2,6-Dinitrotoluene	0.299	0.304	-1.7	92	0.00
52 C	Acenaphthene	1.086	1.091	-0.5	93	0.00
53	3-Nitroaniline	0.292	0.298	-2.1	90	0.00
54 P	2,4-Dinitrophenol	0.122	0.127	-4.1	82	0.00
55	Dibenzofuran	1.657	1.654	0.2	93	0.00
56 P	4-Nitrophenol	0.199	0.190	4.5	83	0.00
57	2,4-Dinitrotoluene	0.397	0.407	-2.5	91	0.00
58	Fluorene	1.330	1.315	1.1	92	0.00
59	2,3,4,6-Tetrachlorophenol	0.306	0.320	-4.6	93	0.00
60	Diethylphthalate	1.338	1.312	1.9	91	0.00
61	4-Chlorophenyl-phenylether	0.653	0.649	0.6	92	0.00
62	4-Nitroaniline	0.306	0.308	-0.7	89	0.00
63	Azobenzene	1.267	1.260	0.6	93	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	91	0.00
65	4,6-Dinitro-2-methylphenol	0.102	0.105	-2.9	87	0.00
66 c	n-Nitrosodiphenylamine	0.591	0.579	2.0	91	0.00
67	4-Bromophenyl-phenylether	0.206	0.203	1.5	93	0.00
68	Hexachlorobenzene	0.238	0.233	2.1	91	0.00
69	Atrazine	0.166	0.168	-1.2	110	0.00
70 C	Pentachlorophenol	0.105	0.110	-4.8	87	0.00
71	Phenanthrene	0.961	0.960	0.1	93	0.00
72	Anthracene	0.940	0.932	0.9	92	0.00
73	Carbazole	0.905	0.906	-0.1	93	0.00
74	Di-n-butylphthalate	1.044	1.047	-0.3	93	0.00
75 C	Fluoranthene	1.044	1.089	-4.3	98	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	102	0.00
77	Benzidine	0.581	0.492	15.3	87	0.00
78	Pyrene	1.849	1.784	3.5	98	0.00
79 S	Terphenyl-d14	1.284	1.204	6.2	96	-0.01
80	Butylbenzylphthalate	0.666	0.675	-1.4	101	0.00
81	Benzo(a)anthracene	1.324	1.309	1.1	104	0.00
82	3,3'-Dichlorobenzidine	0.397	0.378	4.8	94	0.00
83	Chrysene	1.211	1.194	1.4	101	0.00
84	Bis(2-ethylhexyl)phthalate	0.841	0.854	-1.5	105	0.00
85 c	Di-n-octyl phthalate	1.150	1.214	-5.6	109	0.01
86 I	Perylene-d12	1.000	1.000	0.0	100	0.00
87	Indeno(1,2,3-cd)pyrene	1.303	1.273	2.3	98	0.00
88	Benzo(b)fluoranthene	1.256	1.296	-3.2	109	0.00
89	Benzo(k)fluoranthene	1.099	1.007	8.4	92	0.00
90 C	Benzo(a)pyrene	1.021	1.000	2.1	100	0.01
91	Dibenzo(a,h)anthracene	1.067	1.042	2.3	98	0.00
92	Benzo(g,h,i)perylene	1.089	1.078	1.0	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112524\
Data File : BF140590.D
Acq On : 25 Nov 2024 09:33
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 25 18:48:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0