

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF111324

Review By	yogesh	Review On	11/14/2024 3:08:53 AM
Supervise By	mohammad	Supervise On	11/15/2024 5:49:17 AM
SubDirectory	BF111324	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12327,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140331.D	13 Nov 2024 08:35	RC/JU	Ok
2	SSTDICC2.5	BF140332.D	13 Nov 2024 09:01	RC/JU	Ok
3	SSTDICC005	BF140333.D	13 Nov 2024 09:27	RC/JU	Ok,M
4	SSTDICC010	BF140334.D	13 Nov 2024 09:53	RC/JU	Ok,M
5	SSTDICC020	BF140335.D	13 Nov 2024 10:29	RC/JU	Ok
6	SSTDICCC040	BF140336.D	13 Nov 2024 10:55	RC/JU	Not Ok
7	SSTDICC050	BF140337.D	13 Nov 2024 11:21	RC/JU	Ok
8	SSTDICC060	BF140338.D	13 Nov 2024 11:47	RC/JU	Ok
9	SSTDICC080	BF140339.D	13 Nov 2024 12:13	RC/JU	Ok,M
10	SSTDICCC040	BF140340.D	13 Nov 2024 12:48	RC/JU	Ok
11	SSTDICV040	BF140341.D	13 Nov 2024 13:19	RC/JU	Ok
12	PB164401BL	BF140342.D	13 Nov 2024 13:45	RC/JU	Ok
13	P4495-11	BF140343.D	13 Nov 2024 15:25	RC/JU	Dilution
14	P4495-11DL	BF140344.D	13 Nov 2024 16:02	RC/JU	Ok
15	P3845-03	BF140345.D	13 Nov 2024 16:40	RC/JU	Dilution
16	P3845-03DL	BF140346.D	13 Nov 2024 17:06	RC/JU	Ok
17	P3845-06	BF140347.D	13 Nov 2024 17:32	RC/JU	Dilution
18	P3845-06DL	BF140348.D	13 Nov 2024 17:58	RC/JU	Not Ok
19	SP6681	BF140349.D	13 Nov 2024 18:39	RC/JU	Ok,NS

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF111824

Review By	yogesh	Review On	11/19/2024 3:20:55 AM
Supervise By	mohammad	Supervise On	11/19/2024 3:30:54 AM
SubDirectory	BF111824	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12327,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140442.D	18 Nov 2024 09:18	RC/JU	Ok
2	SSTDCCC040	BF140443.D	18 Nov 2024 10:13	RC/JU	Ok,M
3	PB165029BL	BF140444.D	18 Nov 2024 10:39	RC/JU	Ok
4	PB165029BS	BF140445.D	18 Nov 2024 11:05	RC/JU	Ok,M
5	PB164846BS	BF140446.D	18 Nov 2024 11:32	RC/JU	Ok,M
6	PB164846BSD	BF140447.D	18 Nov 2024 11:58	RC/JU	Ok,M
7	PB164940BL	BF140448.D	18 Nov 2024 12:24	RC/JU	Ok
8	PB164940BS	BF140449.D	18 Nov 2024 12:50	RC/JU	Ok,M
9	PB164965BL	BF140450.D	18 Nov 2024 13:16	RC/JU	Ok
10	PB164965BS	BF140451.D	18 Nov 2024 13:43	RC/JU	Ok,M
11	PB164969BS	BF140452.D	18 Nov 2024 14:19	RC/JU	Ok,M
12	PB164969BL	BF140453.D	18 Nov 2024 14:45	RC/JU	Ok
13	PB164880TB	BF140454.D	18 Nov 2024 15:11	RC/JU	Ok
14	DFTPP	BF140455.D	18 Nov 2024 15:48	RC/JU	Ok
15	SSTDCCC040	BF140456.D	18 Nov 2024 16:14	RC/JU	Ok,M
16	PB164944TB	BF140457.D	18 Nov 2024 16:40	RC/JU	Ok
17	P4821-04RE	BF140458.D	18 Nov 2024 17:13	RC/JU	Confirms
18	P4861-01	BF140459.D	18 Nov 2024 17:40	RC/JU	Ok
19	P4871-01	BF140460.D	18 Nov 2024 18:06	RC/JU	Ok
20	P4848-02	BF140461.D	18 Nov 2024 18:32	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF111924

Review By	yogesh	Review On	11/20/2024 7:05:01 AM
Supervise By	mohammad	Supervise On	11/20/2024 7:31:21 AM
SubDirectory	BF111924	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12327,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140462.D	19 Nov 2024 09:32	RC/JU	Ok
2	SSTDCCC040	BF140463.D	19 Nov 2024 10:34	RC/JU	Ok,M
3	PB165019TB	BF140464.D	19 Nov 2024 11:00	RC/JU	Ok
4	P4848-02MS	BF140465.D	19 Nov 2024 11:36	RC/JU	Ok,M
5	P4848-02MSD	BF140466.D	19 Nov 2024 12:02	RC/JU	Ok,M
6	P4870-04	BF140467.D	19 Nov 2024 12:28	RC/JU	Ok
7	P4852-01	BF140468.D	19 Nov 2024 12:54	RC/JU	Ok,M
8	P4852-02	BF140469.D	19 Nov 2024 13:21	RC/JU	Ok,M
9	P4852-03	BF140470.D	19 Nov 2024 13:47	RC/JU	Ok,M
10	P4852-03MS	BF140471.D	19 Nov 2024 14:13	RC/JU	Ok,M
11	P4852-03MSD	BF140472.D	19 Nov 2024 14:39	RC/JU	Ok,M
12	P4771-05	BF140473.D	19 Nov 2024 15:06	RC/JU	Ok,M
13	P4771-07	BF140474.D	19 Nov 2024 15:32	RC/JU	Ok,M
14	P4889-03	BF140475.D	19 Nov 2024 15:58	RC/JU	Ok,M
15	P4869-01	BF140476.D	19 Nov 2024 16:24	RC/JU	Ok
16	P4870-01	BF140477.D	19 Nov 2024 16:51	RC/JU	Ok
17	P4857-03	BF140478.D	19 Nov 2024 17:17	RC/JU	Ok,M
18	P4889-05	BF140479.D	19 Nov 2024 17:43	RC/JU	Dilution
19	P4889-01	BF140480.D	19 Nov 2024 18:10	RC/JU	Ok,M
20	P4849-01	BF140481.D	19 Nov 2024 18:36	RC/JU	Ok,M
21	P4768-03	BF140482.D	19 Nov 2024 19:02	RC/JU	Ok

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF111924

Review By	yogesh	Review On	11/20/2024 7:05:01 AM
Supervise By	mohammad	Supervise On	11/20/2024 7:31:21 AM
SubDirectory	BF111924	HP Acquire Method	BNA_F
		HP Processing Method	Bf111324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12327,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P4859-01RE	BF140483.D	19 Nov 2024 19:28	RC/JU	Confirms
23	P4908-01	BF140484.D	19 Nov 2024 19:54	RC/JU	Ok
24	P4908-03	BF140485.D	19 Nov 2024 20:20	RC/JU	Ok,M
25	P4768-07	BF140486.D	19 Nov 2024 20:47	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/22/2024 4:01:03 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12329,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140526.D	21 Nov 2024 10:17	RC/JU	Ok
2	SSTDCCC040	BF140527.D	21 Nov 2024 10:44	RC/JU	Not Ok
3	SSTDICC2.5	BF140528.D	21 Nov 2024 11:13	RC/JU	Ok
4	SSTDICC005	BF140529.D	21 Nov 2024 11:39	RC/JU	Ok
5	SSTDICC010	BF140530.D	21 Nov 2024 12:05	RC/JU	Ok,M
6	SSTDICC020	BF140531.D	21 Nov 2024 12:32	RC/JU	Ok,M
7	SSTDICCC040	BF140532.D	21 Nov 2024 12:58	RC/JU	Ok,M
8	SSTDICC050	BF140533.D	21 Nov 2024 13:25	RC/JU	Ok
9	SSTDICC060	BF140534.D	21 Nov 2024 13:51	RC/JU	Ok
10	SSTDICC080	BF140535.D	21 Nov 2024 14:18	RC/JU	Ok
11	SSTDICV040	BF140536.D	21 Nov 2024 15:07	RC/JU	Ok,M
12	PB165144BL	BF140537.D	21 Nov 2024 15:34	RC/JU	Ok
13	DFTPP	BF140538.D	21 Nov 2024 16:27	RC/JU	Ok
14	SSTDCCC040	BF140539.D	21 Nov 2024 16:54	RC/JU	Ok,M
15	PB165060TB	BF140540.D	21 Nov 2024 17:20	RC/JU	Ok
16	P4887-06	BF140541.D	21 Nov 2024 17:55	RC/JU	Ok
17	P4887-02	BF140542.D	21 Nov 2024 18:21	RC/JU	Ok
18	P4870-16	BF140543.D	21 Nov 2024 18:48	RC/JU	Ok
19	P4870-15	BF140544.D	21 Nov 2024 19:14	RC/JU	Ok
20	P4870-14	BF140545.D	21 Nov 2024 19:40	RC/JU	Ok
21	P4870-13	BF140546.D	21 Nov 2024 20:07	RC/JU	Ok

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF112124

Review By	yogesh	Review On	11/22/2024 3:53:36 AM
Supervise By	mohammad	Supervise On	11/22/2024 4:01:03 AM
SubDirectory	BF112124	HP Acquire Method	BNA_F
		HP Processing Method	bf112124
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621		
CCC	SP6624		
Internal Standard/PEM	S12329,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P4929-02	BF140547.D	21 Nov 2024 20:33	RC/JU	Ok
23	P4924-04	BF140548.D	21 Nov 2024 20:59	RC/JU	Ok
24	P4921-01	BF140549.D	21 Nov 2024 21:25	RC/JU	Ok
25	P4916-12	BF140550.D	21 Nov 2024 21:51	RC/JU	Ok
26	P4916-08	BF140551.D	21 Nov 2024 22:18	RC/JU	Ok
27	P4916-04	BF140552.D	21 Nov 2024 22:44	RC/JU	Ok
28	P4887-01	BF140553.D	21 Nov 2024 23:10	RC/JU	Ok
29	P4916-09	BF140554.D	21 Nov 2024 23:36	RC/JU	Ok
30	P4916-09MS	BF140555.D	22 Nov 2024 00:02	RC/JU	Ok,NS
31	P4916-09MSD	BF140556.D	22 Nov 2024 00:28	RC/JU	Ok,NS
32	P4916-01	BF140557.D	22 Nov 2024 00:54	RC/JU	Ok
33	P4916-05RE	BF140558.D	22 Nov 2024 01:21	RC/JU	Confirms
34	P4887-05RE	BF140559.D	22 Nov 2024 01:47	RC/JU	Confirms
35	P4924-01	BF140560.D	22 Nov 2024 02:13	RC/JU	ReRun
36	P4936-01	BF140561.D	22 Nov 2024 02:39	RC/JU	Ok,NS
37	P4929-01	BF140562.D	22 Nov 2024 03:06	RC/JU	Ok,NS

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF112524

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF112524	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12330,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF140589.D	25 Nov 2024 09:07	RC/JU	Ok
2	SSTDCCC040	BF140590.D	25 Nov 2024 09:33	RC/JU	Ok,NR
3	PB165123TB	BF140591.D	25 Nov 2024 09:59	RC/JU	Ok
4	PB165155BS	BF140592.D	25 Nov 2024 10:25	RC/JU	Ok,NR
5	PB165155BL	BF140593.D	25 Nov 2024 10:51	RC/JU	Ok
6	PB165086BS	BF140594.D	25 Nov 2024 11:17	RC/JU	Not Ok
7	PB165111TB	BF140595.D	25 Nov 2024 11:43	RC/JU	Ok
8	PB165052BS	BF140596.D	25 Nov 2024 12:09	RC/JU	Ok,NR
9	PB164986TB	BF140597.D	25 Nov 2024 12:36	RC/JU	Ok
10	PB165152BS	BF140598.D	25 Nov 2024 13:02	RC/JU	Ok,NR
11	PB164886TB	BF140599.D	25 Nov 2024 13:27	RC/JU	Ok
12	PB165152BSD	BF140600.D	25 Nov 2024 13:54	RC/JU	Ok,NR
13	PB165152BL	BF140601.D	25 Nov 2024 14:28	RC/JU	Ok
14	PB165185BS	BF140602.D	25 Nov 2024 14:54	RC/JU	Ok,NR
15	DFTPP	BF140603.D	25 Nov 2024 15:23	RC/JU	Ok
16	SSTDCCC040	BF140604.D	25 Nov 2024 15:49	RC/JU	Ok,NR
17	PB165052BL	BF140605.D	25 Nov 2024 16:17	RC/JU	Ok
18	P4947-01	BF140606.D	25 Nov 2024 16:49	RC/JU	Ok
19	P4892-04	BF140607.D	25 Nov 2024 17:15	RC/JU	Ok
20	P4860-09	BF140608.D	25 Nov 2024 17:41	RC/JU	Ok
21	P4860-01	BF140609.D	25 Nov 2024 18:07	RC/JU	Ok

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF112524

Review By		Review On			
Supervise By		Supervise On			
SubDirectory	BF112524	HP Acquire Method	BNA_F	HP Processing Method	bf112124
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6628,SP6627,SP6626,SP6625,SP6624,SP6623,SP6622,SP6621				
CCC	SP6624				
Internal Standard/PEM	S12330,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

22	P4960-01	BF140610.D	25 Nov 2024 18:34	RC/JU	Ok
23	P4960-05	BF140611.D	25 Nov 2024 19:00	RC/JU	Ok
24	P4892-03	BF140612.D	25 Nov 2024 19:26	RC/JU	Ok
25	P4892-03MS	BF140613.D	25 Nov 2024 19:52	RC/JU	Ok,NR
26	P4892-03MSD	BF140614.D	25 Nov 2024 20:18	RC/JU	Ok,NR
27	P4860-06	BF140615.D	25 Nov 2024 20:45	RC/JU	Ok
28	P4860-09MS	BF140616.D	25 Nov 2024 21:11	RC/JU	Not Ok
29	P4860-09MSD	BF140617.D	25 Nov 2024 21:37	RC/JU	Not Ok
30	P4960-03	BF140618.D	25 Nov 2024 22:03	RC/JU	ReRun
31	P4860-10	BF140619.D	25 Nov 2024 22:29	RC/JU	ReRun
32	P4860-04	BF140620.D	25 Nov 2024 22:55	RC/JU	Ok
33	P4860-07	BF140621.D	25 Nov 2024 23:22	RC/JU	Ok
34	P4860-03	BF140622.D	25 Nov 2024 23:48	RC/JU	Ok
35	P4951-01	BF140623.D	26 Nov 2024 00:14	RC/JU	ReRun
36	P4954-01	BF140624.D	26 Nov 2024 00:40	RC/JU	Ok,NR
37	P4954-01MS	BF140625.D	26 Nov 2024 01:06	RC/JU	Ok,NR
38	P4954-01MSD	BF140626.D	26 Nov 2024 01:32	RC/JU	Ok,NR
39	P4954-03	BF140627.D	26 Nov 2024 01:58	RC/JU	ReRun

M : Manual Integration