

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/19/2024

Lab File ID: BF140463.D Time Analyzed: 10:34

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	137275	6.875	502770	8.16	278131	9.91
UPPER LIMIT	274550	7.375	1005540	8.657	556262	10.41
LOWER LIMIT	68637.5	6.375	251385	7.657	139066	9.41
EPA SAMPLE NO.						
01 PB165019TB	144350	6.87	545389	8.15	303191	9.90
02 TP-1MS	148608	6.88	566127	8.15	313272	9.91
03 TP-1MSD	147727	6.88	561678	8.15	311833	9.91

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.