

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871

EPA Sample No.: SSTDCCC040 Date Analyzed: 11/25/2024

Lab File ID: BF140604.D Time Analyzed: 15:49

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	106607	6.869	393427	8.15	218835	9.91
UPPER LIMIT	213214	7.369	786854	8.651	437670	10.41
LOWER LIMIT	53303.5	6.369	196714	7.651	109418	9.41
EPA SAMPLE NO.						
01 PB165052BL	100879	6.87	377922	8.15	213827	9.90

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.