

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: AECO02

Lab Code: CHEM

SAS No.: P4871

SDG NO.: P4871

Lab File ID: BF140462.D

DFTPP Injection Date: 11/19/2024

Instrument ID: BNA_F

DFTPP Injection Time: 09:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	37.8
68	Less than 2.0% of mass 69	0.7 (1.9) 1
69	Mass 69 relative abundance	38.6
70	Less than 2.0% of mass 69	0.2 (0.5) 1
127	10.0 - 80.0% of mass 198	49.9
197	Less than 2.0% of mass 198	0.8
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	27.6
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	14.7
442	Greater than 50% of mass 198	91.9
443	15.0 - 24.0% of mass 442	17 (18.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF140463.D	11/19/2024	10:34
PB165019TB	PB165019TB	BF140464.D	11/19/2024	11:00
TP-1MS	P4848-02MS	BF140465.D	11/19/2024	11:36
TP-1MSD	P4848-02MSD	BF140466.D	11/19/2024	12:02