

Surrogate Summary

SDG No.: P4871

Client: AECOM

Analytical Method: SW8260D

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P4871-01	WC-10-A-202411	1,2-Dichloroethane-d4	50	48.8	98	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	47.6	95	86	113
		4-Bromofluorobenzene	50	55.9	112	77	121
P4871-01DL	WC-10-A-202411DL	1,2-Dichloroethane-d4	50	52.1	104	74	125
		Dibromofluoromethane	50	47.2	94	75	124
		Toluene-d8	50	45.9	92	86	113
		4-Bromofluorobenzene	50	47.9	96	77	121

Surrogate Summary

SDG No.: P4871

Client: AECOM

Analytical Method: SW8260-Low

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
VN1118WBL01	VN1118WBL01	1,2-Dichloroethane-d4	50	52.4	105	74	125
		Dibromofluoromethane	50	49.6	99	75	124
		Toluene-d8	50	45.7	91	86	113
		4-Bromofluorobenzene	50	45.2	90	77	121
VN1118WBS01	VN1118WBS01	1,2-Dichloroethane-d4	50	51.1	102	74	125
		Dibromofluoromethane	50	51.7	103	75	124
		Toluene-d8	50	51.1	102	86	113
		4-Bromofluorobenzene	50	56.6	113	77	121
VN1118WBSD0	VN1118WBSD01	1,2-Dichloroethane-d4	50	47.7	95	74	125
		Dibromofluoromethane	50	47.3	95	75	124
		Toluene-d8	50	46.1	92	86	113
		4-Bromofluorobenzene	50	49.8	100	77	121
VN1119WBL01	VN1119WBL01	1,2-Dichloroethane-d4	50	51.3	103	74	125
		Dibromofluoromethane	50	48.4	97	75	124
		Toluene-d8	50	46.5	93	86	113
		4-Bromofluorobenzene	50	45.9	92	77	121
VN1119WBS01	VN1119WBS01	1,2-Dichloroethane-d4	50	47.8	96	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	45.6	91	86	113
		4-Bromofluorobenzene	50	48.2	96	77	121