

**Laboratory Control Sample/Laboratory Control Sample Duplicate Summary**

**SW-846**

**SDG No.:** P4871

**Client:** AECOM

**Analytical Method:** SW8260-Low

**Datafile :** VN084923.D

| Lab Sample ID | Parameter            | Spike | Result | Unit | Rec | RPD | Qual | Low | Limits<br>High | RPD |
|---------------|----------------------|-------|--------|------|-----|-----|------|-----|----------------|-----|
| VN1118WBSD01  | Vinyl chloride       | 20    | 18.2   | ug/L | 91  | 12  |      | 65  | 117            | 20  |
|               | 1,1-Dichloroethene   | 20    | 18.1   | ug/L | 91  | 13  |      | 74  | 110            | 20  |
|               | 2-Butanone           | 100   | 100    | ug/L | 100 | 18  |      | 65  | 122            | 20  |
|               | Carbon Tetrachloride | 20    | 18.5   | ug/L | 93  | 15  |      | 77  | 113            | 20  |
|               | Chloroform           | 20    | 19.0   | ug/L | 95  | 15  |      | 79  | 113            | 20  |
|               | Benzene              | 20    | 17.9   | ug/L | 90  | 16  |      | 82  | 109            | 20  |
|               | 1,2-Dichloroethane   | 20    | 18.6   | ug/L | 93  | 17  |      | 80  | 115            | 20  |
|               | Trichloroethene      | 20    | 17.5   | ug/L | 88  | 15  |      | 77  | 113            | 20  |
|               | Tetrachloroethene    | 20    | 17.1   | ug/L | 86  | 17  |      | 67  | 123            | 20  |
|               | Chlorobenzene        | 20    | 17.7   | ug/L | 89  | 19  |      | 82  | 109            | 20  |