

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P4871

Client: AECOM

Analytical Method: SW8260-Low

Datafile : VN084938.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1119WBS01	Vinyl chloride	20	19.4	ug/L	97			65	117	
	1,1-Dichloroethene	20	18.9	ug/L	95			74	110	
	2-Butanone	100	120	ug/L	120			65	122	
	Carbon Tetrachloride	20	19.5	ug/L	98			77	113	
	Chloroform	20	20.7	ug/L	104			79	113	
	Benzene	20	19.3	ug/L	97			82	109	
	1,2-Dichloroethane	20	21.1	ug/L	106			80	115	
	Trichloroethene	20	18.7	ug/L	94			77	113	
	Tetrachloroethene	20	19.2	ug/L	96			67	123	
	Chlorobenzene	20	18.9	ug/L	95			82	109	