

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871
Lab File ID: VN084911.D Date Analyzed: 11/18/2024
Instrument ID: MSVOA_N Time Analyzed: 09:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	168638	8.22	277153	9.09	245519	11.87
UPPER LIMIT	337276	8.718	554306	9.594	491038	12.365
LOWER LIMIT	84319	7.718	138577	8.594	122760	11.365
EPA SAMPLE NO.						
WC-10-A-202411	212261	8.22	374690	9.10	341618	11.87
VN1118WBL01	175680	8.22	312901	9.10	269131	11.87
VN1118WBS01	184217	8.22	316400	9.09	286519	11.87
VN1118WBSD01	187181	8.22	326112	9.10	290218	11.87

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.