

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871
Lab File ID: VN084935.D Date Analyzed: 11/19/2024
Instrument ID: MSVOA_N Time Analyzed: 09:27
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	191121	8.22	319795	9.09	272380	11.87
UPPER LIMIT	382242	8.718	639590	9.594	544760	12.365
LOWER LIMIT	95560.5	7.718	159898	8.594	136190	11.365
EPA SAMPLE NO.						
WC-10-A-202411DL	165251	8.22	308690	9.10	269378	11.87
VN1119WBL01	193589	8.22	346862	9.10	301411	11.87
VN1119WBS01	167364	8.22	286809	9.10	252679	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.