

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: AECO02
 Lab Code: CHEM Case No.: P4871 SAS No.: P4871 SDG NO.: P4871
 Lab File ID: VN084911.D Date Analyzed: 11/18/2024
 Instrument ID: MSVOA_N Time Analyzed: 09:51
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	120987	13.788				
UPPER LIMIT	241974	14.288				
LOWER LIMIT	60493.5	13.288				
EPA SAMPLE NO.						
WC-10-A-202411	177257	13.79				
VN1118WBL01	114405	13.79				
VN1118WBS01	143055	13.79				
VN1118WBSD01	145074	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.