

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GENERAL CHEMISTRY CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: P4873

MATRIX: Water

METHOD: 9056A,9060A,SM2320 B

- | | NA | NO | YES |
|---|----|----|-----|
| 1. Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 2. Matrix Spike Duplicate Recoveries Met Criteria | | ✓ | |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| The Blank Spike met requirements for all samples. The Matrix Spike (FSND-MW-EVAL-04D-20241113MS) analysis met criteria for all samples except for Chloride due to Matrix interference. The Matrix Spike (FSND-MW-30-20241114MS) analysis met criteria for all samples except for Chloride due to Matrix interference. The Matrix Spike Duplicate (FSND-MW-EVAL-04D-20241113MSD) analysis met criteria for all samples except for Chloride due to matrix interference. The Matrix Spike Duplicate (FSND-MW-30-20241114MSD) analysis met criteria for all samples except for Chloride due to matrix interference. | | | |
| 3. Sample Duplicate Analysis Met QC Criteria | | | ✓ |
| If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | |
| 4. Digestion Holding Time Met | | | ✓ |
| If not met, list number of days exceeded for each sample: | | | |

ADDITIONAL COMMENTS:

QA REVIEW

Date