

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: EA Engineering Science & Technology Date Collected: 11/14/24 10:55

Project: Scotia, NY - Annual Testing Date Received: 11/15/24

Client Sample ID: FSND-MW-30-20241114 SDG No.: P4873

Lab Sample ID: P4873-02 Matrix: WATER

% Solid: 0

Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Alkalinity	28.4		1	1.00	1.00	2.00	mg/L		11/19/24 12:18	SM 2320 B-11
Chloride	56.5	OR	1	0.011	0.30	0.60	mg/L		11/15/24 14:43	9056A
Nitrate	0.25	U	1	0.0034	0.25	0.50	mg/L		11/15/24 14:43	9056A
Sulfate	1.20	J	1	0.032	1.50	3.00	mg/L		11/15/24 14:43	9056A
TOC	2.30		1	0.19	0.50	1.00	mg/L		11/19/24 01:14	9060A

Comments: The alkalinity to pH 4.34=28.4 mg CaCO3/L

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits