

RETENTION TIMES OF INITIAL CALIBRATION

Contract: EAEN05

Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG NO.: P4873

Instrument ID: FID_A Calibration Date(s): 11/19/2024 11/19/2024

Calibration Times: 10:21 10:57

GC Column: RT-U Plot ID: 0.32 (mm)

LAB FILE ID:	RT 500 = <u>FA004264.D</u>	RT 300 = <u>FA004265.D</u>
RT 200 = <u>FA004266.D</u>	RT 100 = <u>FA004267.D</u>	RT 050 = <u>FA004268.D</u>

COMPOUND	RT 500	RT 300	RT 200	RT 100	RT 050	MEAN RT	RT WINDOW	
							FROM	TO
Ethane	1.74	1.73	1.72	1.71	1.71	1.72	1.62	1.82
Ethylene	1.67	1.66	1.65	1.64	1.65	1.65	1.55	1.75
Methane	1.32	1.31	1.32	1.31	1.32	1.31	1.21	1.41