

CALIBRATION VERIFICATION SUMMARY

Contract: EAEN05

Lab Code: CHEM **Case No.:** P4873 **SAS No.:** P4873 **SDG NO.:** P4873

Continuing Calib Date: 11/19/2024 **Initial Calibration Date(s):** 11/19/2024 11/19/2024

Continuing Calib Time: 13:17 **Initial Calibration Time(s):** 10:21 10:57

GC Column: RT-U Plot **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Ethane	1.69	1.74	1.64	1.84	0.06
Ethylene	1.63	1.67	1.57	1.77	0.04
Methane	1.31	1.32	1.22	1.42	0.01