

Continuing Calibration Summary GASES

Contract: EAEN05 **Initi. Calib. Date(s):** 11/19/2024 11/19/2024
Lab Code: CHEM **Date Analyzed:** 11/19/2024 **Time Analyzed:** 11:33
GC Column: RT-U Plot **ID:** 0.32 (mm) **Data File :** FA004270.D
Client Sample No.: CCAL01 **SAS No.:** P4873 **SDG NO.:** P4873
Lab Sample No.: 200 PPMV GASES CCC **Case No.:** P4873

Analyte	Cal Fac	%D	Flag	Conc PPM	%Rec	Lower Limit	Upper Limit
Ethane	496.35	3		207	103.5	80	120
Ethylene	438.48	3		194	97	80	120
Methane	262.4	4		209	104.5	80	120