

Continuing Calibration Summary GASES

Contract: EAEN05 **Initi. Calib. Date(s):** 11/19/2024 11/19/2024
Lab Code: CHEM **Date Analyzed:** 11/19/2024 **Time Analyzed:** 13:17
GC Column: RT-U Plot **ID:** 0.32 (mm) **Data File :** FA004280.D
Client Sample No.: CCAL02 **SAS No.:** P4873 **SDG NO.:** P4873
Lab Sample No.: 200 PPMV GASES CCC **Case No.:** P4873

Analyte	Cal Fac	%D	Flag	Conc PPM	%Rec	Lower Limit	Upper Limit
Ethane	477.52	1		199	99.5	80	120
Ethylene	388.44	14		171	85.5	80	120
Methane	260.62	4		207	103.5	80	120

Denotes outside control criteria: 30% RSD for initial calibration 20 %D for continuing calibration
(When calibration factor fails correlation coefficient is used as per RSK-175)