

Report of Analysis

Client:	EA Engineering Science & Technology	Date Collected:	11/14/24
Project:	Scotia, NY - Annual Testing	Date Received:	11/15/24
Client Sample ID:	FSND-MW-30-20241114	SDG No.:	P4873
Lab Sample ID:	P4873-02	Matrix:	Water
Analytical Method:	RSK175	% Solid:	0
Water Phase Vol :	36 mL	Decanted:	
Gas Phase Vol :	4 mL	Test:	Gases

File ID/Qc Batch:	Dilution:	Date Analyzed
FA004275.D	1	11/19/24 12:31

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-84-0	Ethane	1.92	U	0.66	1.92	9.60	ug/L
74-85-1	Ethylene	4.30	J	0.93	2.64	13.2	ug/L
74-82-8	Methane	737	E	0.43	0.94	4.70	ug/L

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit