

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: EAEN05

Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG No.: P4873

Instrument ID: MSVOA_X Calibration Date/Time: 11/22/2024 21:05

Lab File ID: VX043975.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.633	0.553		-12.64	50
Chloromethane	0.667	0.604	0.1	-9.44	50
Vinyl Chloride	0.709	0.633		-10.72	50
Bromomethane	0.436	0.468		7.34	50
Chloroethane	0.432	0.412		-4.63	50
Trichlorofluoromethane	1.268	1.357		7.02	50
1,1,2-Trichlorotrifluoroethane	0.605	0.544		-10.08	50
1,1-Dichloroethene	0.576	0.528		-8.33	50
Acetone	0.394	0.292		-25.89	50
Carbon Disulfide	1.188	1.034		-12.96	50
Methyl tert-butyl Ether	2.092	2.090		-0.1	50
Methyl Acetate	0.917	0.846		-7.74	50
Methylene Chloride	0.668	0.624		-6.59	50
trans-1,2-Dichloroethene	0.595	0.569		-4.37	50
1,1-Dichloroethane	1.161	1.155	0.1	-0.52	50
Cyclohexane	0.956	0.884		-7.53	50
2-Butanone	0.508	0.453		-10.83	50
Carbon Tetrachloride	0.500	0.488		-2.4	50
cis-1,2-Dichloroethene	0.735	0.728		-0.95	50
Bromochloromethane	0.511	0.513		0.39	50
Chloroform	1.249	1.246		-0.24	50
1,1,1-Trichloroethane	1.077	1.102		2.32	50
Methylcyclohexane	0.578	0.512		-11.42	50
Benzene	1.387	1.324		-4.54	50
1,2-Dichloroethane	0.557	0.541		-2.87	50
Trichloroethene	0.393	0.317		-19.34	50
1,2-Dichloropropane	0.340	0.329		-3.23	50
Bromodichloromethane	0.482	0.474		-1.66	50
4-Methyl-2-Pentanone	0.537	0.509		-5.21	50
Toluene	0.830	0.802		-3.37	50
t-1,3-Dichloropropene	0.491	0.480		-2.24	50
cis-1,3-Dichloropropene	0.539	0.525		-2.6	50
1,1,2-Trichloroethane	0.343	0.324		-5.54	50
2-Hexanone	0.403	0.376		-6.7	50
Dibromochloromethane	0.344	0.327		-4.94	50
1,2-Dibromoethane	0.346	0.340		-1.73	50
Tetrachloroethene	0.330	0.305		-7.58	50
Chlorobenzene	1.098	1.047	0.3	-4.64	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: EAEN05

Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG No.: P4873

Instrument ID: MSVOA_X Calibration Date/Time: 11/22/2024 21:05

Lab File ID: VX043975.D Init. Calib. Date(s): 11/21/2024 11/21/2024

Heated Purge: (Y/N) N Init. Calib. Time(s): 09:47 12:03

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.861	1.812		-2.63	50
m/p-Xylenes	0.694	0.689		-0.72	50
o-Xylene	0.678	0.676		-0.29	50
Styrene	1.121	1.146		2.23	50
Bromoform	0.240	0.239	0.1	-0.42	50
Isopropylbenzene	3.843	3.785		-1.51	50
1,1,2,2-Tetrachloroethane	1.316	1.298	0.3	-1.37	50
1,3-Dichlorobenzene	1.670	1.571		-5.93	50
1,4-Dichlorobenzene	1.702	1.595		-6.29	50
1,2-Dichlorobenzene	1.685	1.612		-4.33	50
1,2-Dibromo-3-Chloropropane	0.264	0.249		-5.68	50
1,2,4-Trichlorobenzene	0.994	0.936		-5.84	50
1,2,3-Trichlorobenzene	1.006	1.011		0.5	50
1,2-Dichloroethane-d4	0.858	0.889		3.61	50
Dibromofluoromethane	0.357	0.358		0.28	50
Toluene-d8	1.232	1.216		-1.3	50
4-Bromofluorobenzene	0.419	0.417		-0.48	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.