

Report of Analysis

Client:	EA Engineering Science & Technology		Date Collected:	
Project:	Scotia, NY - Annual Testing		Date Received:	
Client Sample ID:	VX1126WBL01		SDG No.:	P4873
Lab Sample ID:	VX1126WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044006.D	1		11/26/24 10:54	VX112624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
75-71-8	Dichlorodifluoromethane	0.50	U	0.21	0.50	1.00	ug/L
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
79-20-9	Methyl Acetate	0.75	U	0.60	0.75	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
110-82-7	Cyclohexane	2.50	U	1.60	2.50	5.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
74-97-5	Bromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L

Report of Analysis

Client:	EA Engineering Science & Technology			Date Collected:	
Project:	Scotia, NY - Annual Testing			Date Received:	
Client Sample ID:	VX1126WBL01			SDG No.:	P4873
Lab Sample ID:	VX1126WBL01			Matrix:	Water
Analytical Method:	SW8260			% Solid:	0
Sample Wt/Vol:	5	Units:	mL	Final Vol:	5000 uL
Soil Aliquot Vol:			uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID :	0.18	Level :	LOW
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044006.D	1		11/26/24 10:54	VX112624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
106-93-4	1,2-Dibromoethane	0.50	U	0.16	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
96-12-8	1,2-Dibromo-3-Chloropropane	0.75	U	0.46	0.75	1.00	ug/L
120-82-1	1,2,4-Trichlorobenzene	0.50	U	0.42	0.50	1.00	ug/L
87-61-6	1,2,3-Trichlorobenzene	0.75	U	0.51	0.75	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.3		81 - 118		101%	SPK: 50
1868-53-7	Dibromofluoromethane	47.7		80 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	50.3		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	46.7		85 - 114		93%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	114000	5.55				
540-36-3	1,4-Difluorobenzene	212000	6.751				
3114-55-4	Chlorobenzene-d5	187000	10.049				
3855-82-1	1,4-Dichlorobenzene-d4	75600	12.018				

Report of Analysis

Client:	EA Engineering Science & Technology		Date Collected:	
Project:	Scotia, NY - Annual Testing		Date Received:	
Client Sample ID:	VX1126WBL01		SDG No.:	P4873
Lab Sample ID:	VX1126WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	DB-624UI	ID : 0.18	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VX044006.D	1		11/26/24 10:54	VX112624

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products