

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
 Data File : VX043952.D
 Acq On : 22 Nov 2024 12:13
 Operator : JC/MD
 Sample : VX1122WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1122WBL01

Quant Time: Nov 22 22:19:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 22 03:45:52 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	114313	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	223688	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192949	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	79805	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99473	50.734	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	101.460%	
35) Dibromofluoromethane	5.379	113	74511	46.617	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	93.240%	
50) Toluene-d8	8.647	98	268134	48.666	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	97.340%	
62) 4-Bromofluorobenzene	11.079	95	89774	47.877	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	95.760%	


Target Compounds Qvalue


 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112224\
Data File : VX043952.D
Acq On : 22 Nov 2024 12:13
Operator : JC/MD
Sample : VX1122WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01


Quant Time: Nov 22 22:19:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112124W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 22 03:45:52 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 731
Delta R.T. 0.006 min
Lab File: VX043952.D
Acq: 22 Nov 2024 12:13

Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01







Tgt Ion:168 Resp: 114313
Ion Ratio Lower Upper
168 100
99 68.5 51.8 77.8



#33
1,2-Dichloroethane-d4
Concen: 50.734 ug/l
RT: 5.952 min Scan# 798
Delta R.T. 0.006 min
Lab File: VX043952.D
Acq: 22 Nov 2024 12:13

Tgt Ion: 65 Resp: 99473
Ion Ratio Lower Upper
65 100
67 50.4 0.0 100.8









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.757 min Scan# 93
 Delta R.T. -0.000 min
 Lab File: VX043952.D
 Acq: 22 Nov 2024 12:13




Instrument :
 MSVOA_X
 ClientSampleId :
 VX1122WBL01


Tgt Ion:	114	Resp:	223688
Ion	Ratio	Lower	Upper
114	100		
63	21.1	0.0	44.0
88	15.8	0.0	32.4



#35
 Dibromofluoromethane
 Concen: 46.617 ug/l
 RT: 5.379 min Scan# 704
 Delta R.T. -0.000 min
 Lab File: VX043952.D
 Acq: 22 Nov 2024 12:13

Tgt Ion:	113	Resp:	74511
Ion	Ratio	Lower	Upper
113	100		
111	100.3	81.0	121.4
192	16.2	13.0	19.6









#50
Toluene-d8
Concen: 48.666 ug/l
RT: 8.647 min Scan# 1239
Delta R.T. 0.006 min
Lab File: VX043952.D
Acq: 22 Nov 2024 12:13




Instrument :
MSVOA_X
ClientSampleId :
VX1122WBL01


Tgt Ion: 98 Resp: 268134
Ion Ratio Lower Upper
98 100
100 63.9 52.6 79.0



#62
4-Bromofluorobenzene
Concen: 47.877 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX043952.D
Acq: 22 Nov 2024 12:13

Tgt Ion: 95 Resp: 89774
Ion Ratio Lower Upper
95 100
174 72.3 0.0 150.4
176 68.7 0.0 146.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1470

Delta R.T. 0.006 min

Lab File: VX043952.D

Acq: 22 Nov 2024 12:13

Instrument :

MSVOA_X

ClientSampleId :

VX1122WBL01





Tgt Ion:117 Resp: 192949

Ion Ratio Lower Upper

117 100

82 55.2 46.4 69.6

119 30.9 24.6 37.0



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.006 min

Lab File: VX043952.D

Acq: 22 Nov 2024 12:13

Tgt Ion:152 Resp: 79805

Ion Ratio Lower Upper

152 100

115 63.1 45.4 136.2

150 153.0 0.0 353.0

