

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
 Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG NO.: P4873
 Lab File ID: VX043950.D Date Analyzed: 11/22/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	119844	5.54	210147	6.75	179277	10.06
UPPER LIMIT	239688	6.038	420294	7.251	358554	10.555
LOWER LIMIT	59922	5.038	105074	6.251	89638.5	9.555
EPA SAMPLE NO.						
FSND-RB-4-20241114	105197	5.55	204646	6.76	177710	10.05
TB-2	112888	5.54	216240	6.76	188871	10.06
VX1122WBL01	114313	5.54	223688	6.76	192949	10.06
VX1122WBS01	122957	5.54	229417	6.76	195535	10.06
VX1122WBSD01	109076	5.54	195872	6.76	171011	10.06

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.