

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG NO.: P4873
Lab File ID: VX044004.D Date Analyzed: 11/26/2024
Instrument ID: MSVOA_X Time Analyzed: 09:59
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	135994	5.54	228315	6.75	197529	10.05
UPPER LIMIT	271988	6.038	456630	7.251	395058	10.549
LOWER LIMIT	67997	5.038	114158	6.251	98764.5	9.549
EPA SAMPLE NO.						
FSND-GEP-3-20241114DL	110646	5.54	210130	6.76	186506	10.06
FSND-MW-28-20241114DL	112571	5.55	214625	6.76	191208	10.06
VX1126WBL01	113627	5.55	212092	6.75	187057	10.05
VX1126WBS01	126484	5.54	228899	6.76	196787	10.06

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.