

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: EAEN05
 Lab Code: CHEM Case No.: P4873 SAS No.: P4873 SDG NO.: P4873
 Lab File ID: VX043950.D Date Analyzed: 11/22/2024
 Instrument ID: MSVOA_X Time Analyzed: 11:22
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	88069	12.018				
UPPER LIMIT	176138	12.518				
LOWER LIMIT	44034.5	11.518				
EPA SAMPLE NO.						
FSND-RB-4-20241114	73808	12.02				
TB-2	77834	12.02				
VX1122WBL01	79805	12.02				
VX1122WBS01	86659	12.02				
VX1122WBSD01	79397	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.