

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51878 MA No.: SDG No.: E29Z5
Instrument ID: CV1 Start Date: 11/26/2024
Analytical Method: CVAA Run Batch: LB133633
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.0	0	0.2	0.22	10	2.5	2.6	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51878 MA No.: SDG No.: E29Z5
Instrument ID: CV1 Start Date: 11/26/2024
Analytical Method: CVAA Run Batch: LB133633
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.9	-1	7.5	7.4	-1	10	10	1