

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011

Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29Z5

Initial Calibration Verification Source : MP83384

Continuing Calibration Verification Source : MP83386

Run Batch: LB133633 Analytical Method: CVAA

Concentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV055				ID: CCV056		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	4.3	106		5.0	4.9	98		4.8	96	

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29Z5
Initial Calibration Verification Source : MP83384
Continuing Calibration Verification Source : MP83386
Run Batch: LB133633 Analytical Method: CVAA
Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV057				ID: CCV058		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury					5.0	4.7	94		4.7	95	