

FORM 4 - IN
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29Z2
 Analytical Method: ICP-MS ICSA Source: EPA-0803
 Instrument ID: P7 ICSB Source: EPA-0803
 Run Batch: LB133596
 Concentration Units: ug/L

Analyte	True		Found			
	ICSA	ICSAB	ICSA	%R	ICSAB	%R
Aluminum	100000	100000	92000	92	92000	92
Antimony	1.5	22	1.2	78	22	98
Arsenic	0.1	19	0.34	340	21	108
Barium	1.2	22	1.4	120	22	100
Beryllium	0	19	0.29		20	106
Cadmium	0.7	20	0.27	39	20	102
Calcium	100000	100000	100000	101	100000	101
Chromium	21	40	21	99	41	103
Cobalt	1	20	1.3	125	22	108
Copper	8	25	7.5	93	27	107
Iron	100000	100000	100000	104	100000	104
Lead	4	25	4.7	117	26	104
Magnesium	100000	100000	96000	96	96000	96
Manganese	7	27	7.9	113	28	103
Nickel	6	24	5.7	95	26	106
Potassium	100000	100000	99000	99	99000	99
Selenium	0.3	19	0	0	20	103
Silver	0	18	0.03		20	109
Sodium	100000	100000	100000	101	100000	101
Thallium	0	21	0.11		22	106
Vanadium	0.5	19	0.19	38	20	106
Zinc	11	29	10	94	30	103