

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	11/20/24 11:06		Kareem	OK
2	S1	S01	CAL2	11/20/24 11:11		Kareem	OK
3	S2	S02	CAL3	11/20/24 11:16		Kareem	OK
4	S3	S03	CAL4	11/20/24 11:20		Kareem	OK
5	S4	S04	CAL5	11/20/24 11:25		Kareem	OK
6	S5	S05	CAL6	11/20/24 11:29		Kareem	OK
7	S6	S06	CAL7	11/20/24 11:33		Kareem	OK
8	ICV006	ICV006	ICV	11/20/24 12:34		Kareem	OK
9	ICB006	ICB006	ICB	11/20/24 12:38		Kareem	OK
10	ICSA006	ICSA006	ICSA	11/20/24 12:52		Kareem	OK
11	ICSAB006	ICSAB006	ICSAB	11/20/24 12:56		Kareem	OK
12	CCV034	CCV034	CCV	11/20/24 13:01		Kareem	OK
13	CCB034	CCB034	CCB	11/20/24 13:05		Kareem	OK
14	P4880-22	MBHH97	SAM	11/20/24 13:09		Kareem	OK
15	P4907-18	MBHHF0	SAM	11/20/24 13:14		Kareem	OK
16	P4907-19	MBHHF1	SAM	11/20/24 13:19		Kareem	OK
17	PB165068BL	PBW068	MB	11/20/24 13:23		Kareem	OK
18	PB165068BS	LCS068	LCS	11/20/24 13:28		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

19	P4879-01	E29X1	SAM	11/20/24 13:32	Ca high	Kareem	Dilution
20	P4879-02	E29X2	SAM	11/20/24 13:37	Ca high	Kareem	Dilution
21	P4879-03	E29X3	SAM	11/20/24 13:41	Ca high	Kareem	Dilution
22	P4879-04	E29X4	SAM	11/20/24 13:46	Ca high	Kareem	Dilution
23	P4879-05	E29X5	SAM	11/20/24 13:50	Ca high	Kareem	Dilution
24	P4879-06	E29X5D	DUP	11/20/24 13:55	Ca high	Kareem	Dilution
25	P4879-05L	E29X5L	SD	11/20/24 14:00	Ca high	Kareem	Dilution
26	P4879-07	E29X5S	MS	11/20/24 14:04		Kareem	OK
27	P4879-08	E29X6	SAM	11/20/24 14:09	Ca high	Kareem	Dilution
28	P4879-09	E29X7	SAM	11/20/24 14:13	Ca high	Kareem	Dilution
29	P4879-10	E29X8	SAM	11/20/24 14:18	Ca high	Kareem	Dilution
30	P4879-11	E29X9	SAM	11/20/24 14:22		Kareem	OK
31	P4879-12	E29Y0	SAM	11/20/24 14:27	Ca high	Kareem	Dilution
32	P4879-13	E29Y1	SAM	11/20/24 14:31	Ca high	Kareem	Dilution
33	P4879-14	E29Y2	SAM	11/20/24 14:36	Ca high	Kareem	Dilution
34	CCV035	CCV035	CCV	11/20/24 14:41		Kareem	OK
35	CCB035	CCB035	CCB	11/20/24 14:45		Kareem	OK
36	P4879-15	E29Y3	SAM	11/20/24 14:50	Ca high	Kareem	Dilution
37	P4879-16	E29Y4	SAM	11/20/24 14:54	Ca high	Kareem	Dilution
38	P4879-17	E29Y5	SAM	11/20/24 14:59	Ca high	Kareem	Dilution

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

39	P4879-18	E29Y9	SAM	11/20/24 15:03	Ca high	Kareem	Dilution
40	P4879-19	E29Z0	SAM	11/20/24 15:08	Ca high	Kareem	Dilution
41	P4879-20	E29Y6	SAM	11/20/24 15:12	Ca high	Kareem	Dilution
42	P4879-21	E29Y7	SAM	11/20/24 15:17		Kareem	OK
43	P4879-22	E2A11	SAM	11/20/24 15:22	Ca high	Kareem	Dilution
44	PB165106BL	PBS106	MB	11/20/24 15:26		Kareem	OK
45	PB165106BS	LCS106	LCS	11/20/24 15:31		Kareem	OK
46	P4880-01	MBHH78	SAM	11/20/24 15:35		Kareem	OK
47	P4880-02	MBHH79	SAM	11/20/24 15:40		Kareem	OK
48	P4880-03	MBHH80	SAM	11/20/24 15:44		Kareem	OK
49	P4880-04	MBHH81	SAM	11/20/24 15:48		Kareem	OK
50	P4880-05	MBHH82	SAM	11/20/24 15:53		Kareem	OK
51	P4880-06	MBHH83	SAM	11/20/24 15:57		Kareem	OK
52	P4880-07	MBHH84	SAM	11/20/24 16:02		Kareem	OK
53	P4880-08	MBHH85	SAM	11/20/24 16:06		Kareem	OK
54	P4880-09	MBHH86	SAM	11/20/24 16:10		Kareem	OK
55	P4880-10	MBHH87	SAM	11/20/24 16:15		Kareem	OK
56	CCV036	CCV036	CCV	11/20/24 16:19		Kareem	OK
57	CCB036	CCB036	CCB	11/20/24 16:24		Kareem	OK
58	P4880-11	MBHH88	SAM	11/20/24 16:28		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

59	P4880-12	MBHH89	SAM	11/20/24 16:33	Ca high	Kareem	Dilution
60	P4880-13	MBHH90	SAM	11/20/24 16:37		Kareem	OK
61	P4880-14	MBHH91	SAM	11/20/24 16:42	NOT USE	Kareem	OK
62	P4880-15	MBHH91D	DUP	11/20/24 16:46	NOT USE	Kareem	OK
63	P4880-14L	MBHH91L	SD	11/20/24 16:50	NOT USE	Kareem	OK
64	P4880-16	MBHH91S	MS	11/20/24 16:55	NOT USE	Kareem	OK
65	P4880-17	MBHH92	SAM	11/20/24 16:59		Kareem	OK
66	P4880-18	MBHH93	SAM	11/20/24 17:03		Kareem	OK
67	P4880-19	MBHH94	SAM	11/20/24 17:08		Kareem	OK
68	P4880-20	MBHH95	SAM	11/20/24 17:12		Kareem	OK
69	P4880-21	MBHH96	SAM	11/20/24 17:17		Kareem	OK
70	PB165050BL	PBS050	MB	11/20/24 17:21		Kareem	OK
71	PB165050BS	LCS050	LCS	11/20/24 17:25		Kareem	OK
72	P4905-01	MBHHA5	SAM	11/20/24 17:30		Kareem	OK
73	P4905-02	MBHHA6	SAM	11/20/24 17:34	NOT USE	Kareem	OK
74	P4905-03	MBHHA6D	DUP	11/20/24 17:39	NOT USE	Kareem	OK
75	P4905-02L	MBHHA6L	SD	11/20/24 17:43	NOT USE	Kareem	OK
76	P4905-04	MBHHA6S	MS	11/20/24 17:48	NOT USE	Kareem	OK
77	P4905-05	MBHHA7	SAM	11/20/24 17:52		Kareem	OK
78	CCV037	CCV037	CCV	11/20/24 17:56		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM

STD. NAME	STD REF.#
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136
ICV Standard	MP83143
CCV Standard	MP83146
ICSA Standard	MP83144,MP83145
CRI Standard	
LCS Standard	
Chk Standard	MP83148,MP83149

79	CCB037	CCB037	CCB	11/20/24 18:01		Kareem	OK
80	P4905-06	MBHHA8	SAM	11/20/24 18:05		Kareem	OK
81	P4905-07	MBHHA9	SAM	11/20/24 18:10		Kareem	OK
82	P4905-08	MBHHA0	SAM	11/20/24 18:14		Kareem	OK
83	P4905-09	MBHHA1	SAM	11/20/24 18:18		Kareem	OK
84	P4905-10	MBHHC5	SAM	11/20/24 18:23		Kareem	OK
85	P4905-11	MBHHC6	SAM	11/20/24 18:27		Kareem	OK
86	P4905-12	MBHHC7	SAM	11/20/24 18:31		Kareem	OK
87	P4905-13	MBHHC8	SAM	11/20/24 18:36		Kareem	OK
88	P4905-14	MBHHC9	SAM	11/20/24 18:40		Kareem	OK
89	P4905-15	MBHHD0	SAM	11/20/24 18:45		Kareem	OK
90	P4905-16	MBHHD1	SAM	11/20/24 18:49		Kareem	OK
91	P4905-17	MBHHD2	SAM	11/20/24 18:53		Kareem	OK
92	P4905-18	MBHHD3	SAM	11/20/24 18:58		Kareem	OK
93	P4905-19	MBHHD4	SAM	11/20/24 19:02		Kareem	OK
94	P4905-20	MBHHD5	SAM	11/20/24 19:07		Kareem	OK
95	PB165101BL	PBS101	MB	11/20/24 19:11		Kareem	OK
96	PB165101BS	LCS101	LCS	11/20/24 19:15		Kareem	OK
97	P4728-19DL	MBH9F6	SAM	11/20/24 19:20	2x for Ca	Kareem	Confirms
98	P4879-01DL	E29X1	SAM	11/20/24 19:25	2x for Ca	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

99	P4879-02DL	E29X2	SAM	11/20/24 19:29	5x for Ca	Kareem	Confirms
100	CCV038	CCV038	CCV	11/20/24 19:34		Kareem	OK
101	CCB038	CCB038	CCB	11/20/24 19:38		Kareem	OK
102	P4879-03DL	E29X3	SAM	11/20/24 19:43	2x for Ca	Kareem	Confirms
103	P4879-04DL	E29X4	SAM	11/20/24 19:47	5x for Ca	Kareem	Confirms
104	P4879-05DL	E29X5	SAM	11/20/24 19:52	2x for Ca	Kareem	Confirms
105	P4879-06DL	E29X5D	DUP	11/20/24 19:56	2x for Ca	Kareem	Confirms
106	P4879-05LDL	E29X5L	SD	11/20/24 20:01	10x for Ca	Kareem	Confirms
107	P4879-07DL	E29X5S	MS	11/20/24 20:05	Not required	Kareem	Not Ok
108	P4879-08DL	E29X6	SAM	11/20/24 20:10	5x for Ca	Kareem	Confirms
109	P4879-09DL	E29X7	SAM	11/20/24 20:14	5x for Ca	Kareem	Confirms
110	P4879-10DL	E29X8	SAM	11/20/24 20:18	2x for Ca	Kareem	Confirms
111	P4879-12DL	E29Y0	SAM	11/20/24 20:23	2x for Ca	Kareem	Confirms
112	P4879-13DL	E29Y1	SAM	11/20/24 20:27	2x for Ca	Kareem	Confirms
113	P4879-14DL	E29Y2	SAM	11/20/24 20:32	2x for Ca	Kareem	Confirms
114	P4879-15DL	E29Y3	SAM	11/20/24 20:36	2x for Ca	Kareem	Confirms
115	P4879-16DL	E29Y4	SAM	11/20/24 20:41	2x for Ca	Kareem	Confirms
116	P4879-17DL	E29Y5	SAM	11/20/24 20:45	2x for Ca	Kareem	Confirms
117	P4879-18DL	E29Y9	SAM	11/20/24 20:50	2x for Ca	Kareem	Confirms
118	P4879-19DL	E29Z0	SAM	11/20/24 20:54	2x for Ca	Kareem	Confirms

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB133543

Review By	Sarabjit Jaswal	Review On	11/21/2024 7:23:59 AM
Supervise By	Mohan Bera	Supervise On	11/21/2024 7:25:19 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83134,MP83142,MP83140,MP83139,MP83138,MP83137,MP83136		
ICV Standard	MP83143		
CCV Standard	MP83146		
ICSA Standard	MP83144,MP83145		
CRI Standard			
LCS Standard			
Chk Standard	MP83148,MP83149		

119	P4879-20DL	E29Y6	SAM	11/20/24 20:59	2x for Ca	Kareem	Confirms
120	P4879-22DL	E2A11	SAM	11/20/24 21:04	2x for Ca	Kareem	Confirms
121	CCV039	CCV039	CCV	11/20/24 21:08		Kareem	OK
122	CCB039	CCB039	CCB	11/20/24 21:12		Kareem	OK