

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29X3
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133543 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV006				ID: CCV034				ID: CCV035		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	97	1	400000	430000	107	3	410000	103	0
Calcium	10000	10000	103	0	400000	420000	104	3	400000	101	0
Iron	10000	11000	108	0	400000	430000	108	3	410000	104	0
Magnesium	6000	5500	92	0	400000	420000	104	3	400000	101	0
Potassium	9900	9300	94	1	150000	160000	104	3	150000	101	0
Sodium	10000	9900	99	2	400000	430000	107	3	410000	103	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29X3
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133543 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV036				ID: CCV037		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	0	400000	99	0
Calcium					400000	390000	98	0	390000	97	0
Iron					400000	400000	101	0	400000	100	0
Magnesium					400000	400000	99	0	390000	97	0
Potassium					150000	150000	98	0	150000	97	0
Sodium					400000	400000	101	0	400000	99	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51878 MA No. : SDG No.: E29X3
 Initial Calibration Verification Source : MP83143
 Continuing Calibration Verification Source : MP83146
 Run Batch: LB133543 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV038				ID: CCV039		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	101	0	400000	101	0
Calcium					400000	390000	98	0	390000	97	0
Iron					400000	410000	101	0	400000	101	0
Magnesium					400000	390000	98	0	390000	98	0
Potassium					150000	150000	99	0	150000	99	0
Sodium					400000	400000	101	0	400000	100	0