

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51878 MA No.: SDG No.: E29X3
Instrument ID: CV1 Start Date: 11/27/2024
Analytical Method: CVAA Run Batch: LB133670
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	0	0.016	0	0.2	0.21	3	2.5	2.5	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51878 MA No.: SDG No.: E29X3
Instrument ID: CV1 Start Date: 11/27/2024
Analytical Method: CVAA Run Batch: LB133670
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Mercury	5	4.8	-3	7.5	7.6	1	10	10	0