
From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Monday, November 25, 2024 17:04
To: Sohil Jodhani; Nimisha Pandya; Mohammad Ahmed
Cc: REGION_2 (Kumar.Narendra@epa.gov); Region_2 (Brandon-Bazile.Kim@epa.gov); Leung, Christina; Region_2 (Feranda.Jennifer@epa.gov); REGION_2 (Arnone.Russell@epa.gov); Patel, Bhavita; Ringel, Donna
Subject: REGION 2 | CASE 51837 | LAB ACE | SDG BG3J7 & BG3P6 & BG3G8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT | FINAL
Attachments: P4883-04(BG3K1).pdf; P4883-05(BG3K2).pdf; P4883-06(BG3K3).pdf; P4883-16(BG3L5).pdf; P4883-17(BG3L6).pdf; P4885-05(BG3Q1).pdf; P4885-06(BG3Q2).pdf; P4885-07(BG3Q3).pdf; P4885-08(BG3Q4).pdf; P4885-09(BG3Q5).pdf; P4885-10(BG3Q6).pdf; P4885-12(BG3Q8).pdf; P4885-02(BG3P7).pdf; P4882-14(BG3J1).pdf; P4882-15(BG3J2).pdf

EXTERNAL EMAIL - This email was sent by a person from outside your organization. Exercise caution when clicking links, opening attachments or taking further action, before validating its authenticity.

Secured by Check Point

Good afternoon,

Please refer to the Region's response below regarding Case 51837/ SDGs BG3J7, BG3P6 & BG3G8 . Also, please be sure to provide detailed information about these issues in the SDG Narrative and let me know if you have any questions.

ISSUE:

"Lab is sending this email with regards to case 51837.

This is an ongoing case where Lab has received samples for TVOA analysis and samples are found positive with high concentration of target analytes as required dilution analysis. Lab has analyzed samples BG3K1, BG3K2, BG3K3, BG3L5 & BG3L6 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3J1 & BG3J2 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3Q1, BG3Q2, BG3Q3, BG3Q4, BG3Q5, BG3Q6, BG3Q8 & BG3P7 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Please see attached.”

REGION: “Please advise the laboratory to proceed and detail the deviations from SOW directions in their SDG narrative. Laboratory should insert instrument blanks if screening indicates so. Thanks.”

***** Please note that in order to waive/remove defects associated to this issue, please send a detailed waiver request to the appropriate ASB PM.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

GENERAL DYNAMICS
Information Technology

Leave Alert: December 4th and 5th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Kumar, Narendra <Kumar.Narendra@epa.gov>
Sent: Monday, November 25, 2024 4:55 PM
To: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Cc: Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>; Patel, Bhavita <Patel.Bhavita@epa.gov>; DASsupport <DASsupport@gdit.com>; Ringel, Donna <Ringel.Donna@epa.gov>
Subject: RE: REGION 2 | CASE 51837 | LAB ACE | SDG BG3J7 & BG3P6 & BG3G8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello Prakriti,
Please advise the laboratory to proceed and detail the deviations from SOW directions in their SDG narrative. Laboratory should insert instrument blanks if screening indicates so. Thanks.

Narendra Kumar

From: Panday, Prakriti P <Prakriti.Panday@gdit.com>
Sent: Monday, November 25, 2024 4:30 PM

To: Kumar, Narendra <Kumar.Narendra@epa.gov>; Brandon-Bazile, Kim <Brandon-Bazile.Kim@epa.gov>; Leung, Christina (she/her/hers) <Leung.Christina@epa.gov>; Feranda, Jennifer <Feranda.Jennifer@epa.gov>; Arnone, Russell <Arnone.Russell@epa.gov>

Subject: REGION 2 | CASE 51837 | LAB ACE | SDG BG3J7 & BG3P6 & BG3G8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Caution: This email originated from outside EPA, please exercise additional caution when deciding whether to open attachments or click on provided links.

Good afternoon,

Please advise if the Region agrees with the laboratory's proposal for Case 51837/SDGs BG3J7, BG3P6 & BG3G8

ISSUE:

Lab is sending this email with regards to case 51837.

This is an ongoing case where Lab has received samples for TVOA analysis and samples are found positive with high concentration of target analytes as required dilution analysis. Lab has analyzed samples BG3K1, BG3K2, BG3K3, BG3L5 & BG3L6 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3J1 & BG3J2 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3Q1, BG3Q2, BG3Q3, BG3Q4, BG3Q5, BG3Q6, BG3Q8 & BG3P7 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Please see attached.

Thank you,

PRAKRITI PANDAY
Associate Environmental Analyst
571-533-3261
GDIT Federal Civilian Division
Under Contract to EPA
15036 Conference Center Drive
Chantilly, VA 20151
www.gdit.com

Leave Alert: December 4th and 5th

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: DASsupport

Sent: Monday, November 25, 2024 4:16 PM

To: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 2 | CASE 51837 | LAB ACE | SDG BG3J7 & BG3P6 & BG3G8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

Thank you for contacting DAS Support. This is to confirm that your email has been received and is currently being addressed. An updated response will be provided as soon as it is available.

Thank you,

The DAS Support Team

DASsupport@gdit.com

GDIT Federal Civilian Division

Under Contract to EPA

15036 Conference Center Drive

Chantilly, VA 20151

www.gdit.com

This electronic message transmission contains information from GDIT that may be attorney-client privileged, proprietary or confidential. The information in this message is intended only for use by the individual(s) to whom it is addressed. If you believe you have received this message in error, please contact me immediately and be aware that any use, disclosure, copying or distribution of the contents of this message is strictly prohibited. NOTE: Regardless of content, this email shall not operate to bind GDIT to any order or other contract unless pursuant to explicit written agreement or government initiative expressly permitting the use of email for such purpose.

From: Sohil Jodhani <Sohil.Jodhani@alliancetg.com>

Sent: Monday, November 25, 2024 4:07 PM

To: DASsupport <DASsupport@gdit.com>

Cc: Mohammad Ahmed <mohammad.ahmed@alliancetg.com>; Nimisha Pandya <Nimisha.Pandya@AllianceTG.com>

Subject: REGION 2 | CASE 51837 | LAB ACE | SDG BG3J7 & BG3P6 & BG3G8 | SOW SFAM-ORG | CONTRACT 68HERH20D0011 | ISSUE DAS SUPPORT

This Message Is From an External Sender

Please use caution with links, attachments, and any requests for credentials.

Hello,

Lab is sending this email with regards to case 51837.

This is an ongoing case where Lab has received samples for TVOA analysis and samples are found positive with high concentration of target analytes as required dilution analysis. Lab has analyzed samples BG3K1, BG3K2, BG3K3, BG3L5 & BG3L6 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3J1 & BG3J2 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Lab has analyzed samples BG3Q1, BG3Q2, BG3Q3, BG3Q4, BG3Q5, BG3Q6, BG3Q8 & BG3P7 in a continuous analytical sequence. All samples are required dilution to bring target analytes within calibration range. In this case, instrument blank was not analyzed in between the samples due to continuous analytical sequence therefore lab would like to confirm that lab will report undiluted TVOA analysis without instrument blank in between the samples and further dilution analysis in final electronic deliverables. Please note that there is not any other Lab QC failure associated to this analysis.

Please see attached.

Thanks & Regards,



Sohil Jodhani
QA/QC Director
An Alliance Technical Group Company
Main: 908-789-8900
Direct: 908-728-3152
Address: 284 Sheffield St, Ste 1, Mountainside, NJ 07092
www.alliancetg.com

